



MODERN SHOCKPROOF X-RAY AND ELECTROMEDICAL EQUIPMENT


*Distinguished for
Outstanding Performance
and Great Durability*

SIMPLIFIED MANUAL

OF

X-RAY TECHNIC

H. G. FISCHER & CO.



FIVE IMPORTANT FEATURES
of
H. G. FISCHER & CO.
PRODUCTS . . .

- 1 EFFICIENCY
- 2 DURABILITY
- 3 SERVICE
- 4 ECONOMY
- 5 ATTRACTIVENESS



H. G. FISCHER & CO.

Manufacturers

X-ray and Electromedical Equipment

9451-91 W. BELMONT AVENUE

FRANKLIN PARK, ILLINOIS

(SUBURB OF CHICAGO)

SIMPLIFIED MANUAL

OF

X-RAY TECHNIC

FOREWORD

This manual of Medical x-ray technic and related data has been prepared by the X-ray Field Service Department of H. G. FISCHER & CO. and is presented to you with their compliments.

It is not intended to be a complete treatise on radiographic technic, but consists of an accumulation of papers or sets of papers containing general practical information on the subject.

The entire purpose of the manual is to give you, in condensed form, material of a nature that we believe will have the effect of improving radiographic end results and consequently the accuracy of a diagnosis.

Special attention has been paid to positioning and we hope that the positioning information, both written and photographic, will be of value to you.

H. G. FISCHER & CO. are vitally interested in every Fischer installation and you are cordially invited to call upon them at any time for helpful technical suggestions.

H. G. FISCHER & CO.

RECOMMENDED PROCEDURE FOR ADJUSTING AND
SETTING UP EQUIPMENT FOR RADIOGRAPHIC WORK

Experience has proven that if a routine procedure is followed when making a radiograph, many mistakes and omissions will be eliminated with the resulting economy in film.

While experienced operators will have their own ideas of radiographic procedure, it is recommended that operators with little or no experience follow the step by step routine outlined below.

1. After determining what anatomical part is to be radiographed, refer to the technic chart furnished, which shows all the technical factors that will be used, such as milliamperes, kilovoltage, distance and time. The chart will also show if the Bucky diaphragm is to be used. All factors that can be preset should be adjusted before the patient is placed on the table.
2. If high milliamperage values are to be used (50 MA or more) the tube should be prewarmed. Prewarming of the tube before using high milliamperage values will increase tube life. A machine setting of 5 milliamperes, 85 kilovolts, and 10 second time is sufficient to prewarm the x-ray tube. If the x-ray tube has been used within the past two hours, prewarming of the x-ray tube is not necessary.
3. If the machine has an automatic control, place milliamperage selector switch on milliamperage value shown on technic chart. If the control is of the manual type, set milliamperes by means of the main and auxiliary filament controls.
4. If control has a line compensator, adjust line compensator switch so that line compensator meter reads on red line.
5. If bucky is to be used, place timer switch in bucky timer position. If bucky is not used, place timer switch in timer position. Adjust timer for time factor shown on chart.
6. If bucky is used set timing device on Bucky diaphragm for approximately 20% more time than the actual time of exposure. Cock bucky by pulling Cocking handle all the way out.
7. If Bucky is used, make certain that the x-ray tube is centered on the table.
8. Adjust x-ray tube for correct height as indicated on the chart.
9. Place patient on the x-ray table making certain that the patient is centered on the table.
10. Measure patient with centimeter thickness caliper and set kilovoltage according to the chart.
11. Place cassette in bucky tray and position bucky so that part to be radiographed is centered on the cassette.

12. Center x-ray tube on cassette in Bucky diaphragm.
13. If compression band is to be used, place compression band in place and tighten lightly.
14. Double check all procedure such as line compensator, milliamperage selector, kilovoltage, timer and timer switch, bucky timing, bucky cocking and height.
15. Tighten compression band if used.
16. Press timer push button. If Bucky diaphragm is used, it is only necessary to depress timer button for about 1/4 second. If bucky is not used and timer switch is set in timer position, it is necessary to hold timer push button down until timer automatically cuts off.

SPECIAL NOTES

1. Always prewarm cold tube when using high MA values.
2. On manual controls when using high MA values, always test milliamperage setting at a kilovoltage 20 KV below the voltage value being used.
3. Longer tube life will follow if machine is turned off while positioning patient or changing patients.
4. Always use as small a cone as possible and still cover anatomical area to be radiographed.
5. Grasp compression band with hand and release slowly to eliminate injury to patient by the sudden release of the pressure.
6. Always keep loaded cassette in protective film chest until ready to make radiograph.
7. If red position of heat indicator shows on tube, further use must stop until tube cools off and heat indicator goes down.
8. Testing of MA values on manual controls should be done as rapidly as possible as time consumed in testing is the same as in actual exposure.
9. Consult tube loading chart for safe operation of tube.

FLUOROSCOPY

The Following Fluoroscopic Precautions and Procedures Shall Be Observed.

The operator shall allow his eyes to become fully accommodated to the darkness before starting a fluoroscopic examination. This may be done by remaining in a dark room or wearing fluoroscopic goggles provided with special colored lenses. At least five to ten minutes are necessary for proper accommodation of the eyes.

Failure to observe this rule will result in unsatisfactory fluoroscopic vision and will tempt the operator to increase the milliamperage and kilovoltage, beyond the recommended maximum value thus exposing the patient, and under certain conditions himself, to dangerous amounts of radiation.

This rule applies with particular emphasis to the use of head or hand type fluoroscopes.

X-Ray protection devices such as lead rubber gloves, lead rubber aprons, lead protected fluoroscopic chair, etc. shall always be used. The operator's bare hands shall never be exposed to the x-ray beam.

The distance between the target of the x-ray tube and the patient's skin shall be at least ten (10) inches. Greater distances are preferable since they reduce the likelihood of overexposure and improve the fluoroscopic image. The intensity of the x-radiation reaching the patient's skin varies as the square of the distance from the target of the x-ray tube. For example the intensity at ten (10) inch distance is four times that at twenty (20) inches, all other factors remaining the same.

The best practice is to use a minimum of one millimeter aluminum filter or its equivalent between the x-ray tube and the patient, to eliminate long wave length radiation which is injurious to the skin.

The milliamperage setting shall be as low as possible without sacrificing satisfactory fluoroscopic vision. Milliamperage shall in no case exceed five (5) milliamperes.

Kilovoltage is set in accordance with the thickness and density of the body part being examined. For best vision it shall be kept as low as possible and although 100 KV is available in some installations, it is a rare instance that requires more than 85 KV. It is common practice to use not more than five (5) milliamperes at 85 KV for abdominal fluoroscopy. Five (5) milliamperes at 75 to 85 KV for chest, five (5) milliamperes at 60 KV for extremities.

Make short intermittent exposures rather than continued ones. This practice allows the eyes to re-accommodate during the dark intervals and increases the visibility of the image. It reduces the amount of radiation the patient receives and prolongs the life of the x-ray tube.

Restrict the size of the radiation aperture or shutter to the smallest area necessary to cover the part being examined. This prevents needless exposure of other parts of the patient's body and improves the sharpness of the fluoroscopic image.

Consult the tube rating chart for maximum permissible time, kilovoltage and milliamperage which apply to your fluoroscopic tube. Never exceed these ratings.

As a guide in determining the safe exposure factors consult the enclosed table of maximum permissible exposure value in milliamperere seconds. Milliamperere seconds are the product of milliamperes and total actual exposure time in seconds. In using this table the maximum permissible exposure time in seconds is obtained by dividing the milliamperes used into the value of milliamperere seconds permissible, for example, if the milliamperere seconds used are five and the permissible exposure value in milliamperere seconds total 1000 at a given filtration and distance, the milliamperage five (5) is divided into the milliamperage seconds 1000 which gives the time 200 seconds. This figure constitutes the maximum exposure time permissible.

MAXIMUM EXPOSURE FACTORS PERMISSIBLE FOR ALL PARTS
OF THE BODY EXCEPT HEAD WHEN USING 85 KILOVOLTS PEAK

FACTORS APPLY TO BOTH RADIOGRAPHY AND FLUCROSCOPY

Anode Skin Distance Inches	Aluminum filter .5 MM inherent .5 MM external Permissible Milliampere second Exposure
10" -----	510 MAS
12" -----	730 MAS
14" -----	1000 MAS
16" -----	1300 MAS
18" -----	1650 MAS
20" -----	2050 MAS
22" -----	2450 MAS
24" -----	2900 MAS
30" -----	4500 MAS
36" -----	6500 MAS
42" -----	8850 MAS
48" -----	11600 MAS
60" -----	18000 MAS
72" -----	26000 MAS

CAUTION When radiographing the head, reduce above milliampere second factors 30%.

VARIATION IN MAXIMUM PERMISSIBLE EXPOSURE VALUES WITH CHANGE IN KILOVOLTAGE

Kilovolts	Change in permissible exposure values
100 KVP	Reduce by 25%
90 KVP	Reduce by 8%
85 KVP	No change
80 KVP	Increase by 10%
70 KVP	Increase by 35%
60 KVP	Increase by 80%

NOTE Milliampere seconds - Milliamperes X Time in seconds.

Example: 100 Milliamperes 1 second = 100 MAS
 50 Milliamperes 2 seconds = 100 MAS
 25 Milliamperes 4 seconds = 100 MAS
 10 Milliamperes 10 seconds = 100 MAS
 5 Milliamperes 20 seconds = 100 MAS

GENERAL DARK ROOM INFORMATION

Too much stress cannot be laid on the importance of the processing room construction and technical procedures, whether it be for private practice or a large institution. It is to be realized that the quality of the end result depends entirely on the facilities and the procedure used.

No radiograph is complete until it has passed through the developing process. Experience proves that even with the best radiographic equipment and accessories, radiographic technic can be made mediocre by faulty processing room service. Yet it is doubtful if ten percent of film processing laboratories the world over are on a par with the rest of the x-ray department in point of facilities to handle the work. Most laboratories are handicapped because the equipment in the processing room and the haphazard procedures used are inadequate.

LOCATION OF PROCESSING ROOM

An important consideration when planning a dark room is its location. First it must be within a reasonable distance from the x-ray machine. Second, it must be convenient to plumbing and electrical connections. Another important factor is ventilation.

CONSTRUCTION OF PROCESSING ROOM

Construction of processing room with reference to materials used is limited. It must be of such material as to preclude leakage of ordinary light and x-ray if unexposed films are to be stored in the dark room. The size of the dark room is determined by the nature and average amount of work that is to be handled. The minimum dimension should still allow room for convenient and efficient processing procedures. If considerable work is to be handled mazes are generally used which permits entering and leaving the dark room without the necessity of opening and closing doors.

COLOR TREATMENTS OF WALLS

In respect to the color treatment of the walls of the processing room, it is apparent there is much misunderstanding. Most x-ray laboratories utilize either black or some other dark pigment which is not necessary. The problem of color selection for the dark room is not important if the dark room is light proof. If a maze is used in conjunction with the dark room, the maze should be painted a dull dark color so that daylight or white light will be absorbed before it enters the dark room.

In many cases an acid and alkali proof paint is used around the developing tank. This should be a paint which is unaffected by acid or alkali and can be readily washed.

ACCESSORY EQUIPMENT

The following equipment is necessary in the dark room.

1. Safe lights.
2. Storage shelves or cabinets.
3. Loading bench.
4. Racks for developing hangers.
5. Developing tank.
6. Water temperature control, either manual or automatic.

General Dark Room Information -2

7. Thermometer. Be sure to use a thermometer that is designed for developing purposes and which is accurate around 70 degrees Fahrenheit.
8. Developing timer.
9. Drying equipment, either film dryers or racks to hold the films while they are drying.
10. In large installations, a pass box between the radiographic room and the dark room.

ILLUMINATION

Safe lights cause a large percentage of processing room trouble. Fogging of the x-ray film is a phenomenon which the average operator may find difficult to trace to its source, that is to determine whether it is caused by the safelight, x-ray, or a chemical reaction. It is well to remember that there is no such thing as a safe light in the strict sense of the word. All safelights will fog sensitized emulsions by excessive exposure. Thus the problem of film exposure to light is an important one.

In the final analysis, the maximum amount of light intensity or wattage that may be used in the safelight is determined by the following four factors:

1. Location of safelight. That is, safelight to film distance.
2. Safelight filtration or the quality of light allowed to pass through the light filter.
3. Speed of placing the film in hanger.
4. Speed of operation while observing stage of development and rinsing film and placing it in the fixer or hypo solution.

It is apparent from the foregoing, the intensity or wattage of the bulbs used in dark room safelights is dependent upon the speed of the operator when processing a film.

STORAGE FACILITIES

Storage shelves should be of such size as to allow the housing of all sizes of films, cassettes and supplies. If a storage cabinet is used where the covers of the film box are removed, the storage cabinet should be equipped with limit or safety switches which will automatically put out all lights in the dark room with the exception of the safelights, should the storage chest or box be opened accidentally when the white lights are on.

LOADING FACILITIES

The loading bench should be of size which would allow the opening of film packages and loading of cassettes conveniently. The loading bench should always be kept clean as any form of contamination may not only ruin the film to a point where diagnosis is impossible, but also damage the intensifying screen. The loading bench should be situated away from the developing tank so that the developer and hypo will not be accidentally spilled on the screens.

FILM HANGERS

Film hangers are an important item in the processing room procedure. They should be stored in a convenient location where they can be readily reached for loading of films for developing. They must be kept clean and should be thoroughly washed. This can be accomplished while washing the film provided the water in the washing tank covers the cross bar on the hanger. The flexible arms at the top of the developing frame should be adjusted so that the film is held taut in the

General Dark Room Information - 3

frame, thus eliminating a possibility of the film touching an adjacent film or the side of the tank which will cause an undeveloped area.

TANKS

Tanks for developing, rinsing and fixing should be of a size and capacity to meet the requirements of the individual laboratory. It should be capable of maintaining temperatures and be durable with a minimum reaction to chemicals. Many tank combinations are available the sizes and combinations depending entirely upon the amount of work to be done.

WATER TEMPERATURE CONTROL

Inasmuch as the temperature of the developing and fixing solutions are controlled by the rinse and washing water in the water compartment, it is necessary that some control be available for controlling the temperature of the water. In some installations it is merely necessary to have a hot and cold water valve which will permit adjusting the temperature of the water to the correct value. Other installations use thermostatic mixing valves whereby the temperature of the water is automatically controlled. In hot humid climates, very often it is necessary to use a refrigerated tank.

THERMOMETER

The dark room thermometer should be large enough so that its calibration can be read with accuracy when only the dark room lights are on. The dark room thermometer should not be housed in wood for this not only causes error in readings but also causes it to absorb chemicals. Over a period of time this will cause contamination of the solution. To avoid scratches the thermometer should not be allowed to come in contact with the films.

TIMER

A reliable timer is an indispensable item for accurate timing the intervals required in film processing. It should be sturdy and accurate and with a range sufficient to allow the maximum developing time and with a means of signaling when the interval is concluded.

DRYING EQUIPMENT

There are many forms of drying equipment and the type selected will of course depend on the amount of work that the laboratory is doing. Dryers can be obtained which are just ordinary racks where the film is allowed to drain and dry naturally. Other film dryers use forced draft and in some cases heat in order to dry the films quickly. Film dryers are available which not only can be used as film dryers but also as loading benches.

PASS BOX

In large laboratories the use of a pass box between the dark room and the radiographic room is a great convenience. Pass boxes are so designed that if the door on the radiographic room is open the door on the dark room side cannot be opened. This eliminates the possibility of white light entering the darkroom accidentally.

LOADING OF FILMS

When removing a film for loading in a cardboard holder the black paper which encloses the film should not be removed but placed with the film in the holder; this minimizes the possibility of fogging the film. When loading a cassette, however, the operator must be sure to let the paper fall away from the film. This is accomplished by taking the film out with the right hand, separating the film from the paper with the left hand and then allowing the paper to fall away from the film, thus minimizing friction and static. If there is difficulty in replacing the cover on the film carton, the sides should be pressed inward. This releases the air and allows the cover to fall into place.

DEVELOPING AND FIXING

The developing and fixing solution unless used at par strength and according to prescribed time and temperature are possible causes of more dark room trouble than any other factors. Temperature is very important in developing an x-ray film. As the temperature is raised the alkali works faster and as the temperature lowers it is retarded. The operator should consult the recommended developing time which will be found on the label of the developing container. It is also recommended that the developer manufactured by the film manufacturer be used.

Many operators make the mistake of trying to stretch out the life of their developer. This usually will cause a poor grade of radiograph with a decided falling off of the diagnostic qualities, and in the end it is poor economy as films are bound to be ruined and inasmuch as the developer is much cheaper than films it is advantageous to discard the developer before it reaches the point where it will affect the quality of the film.

THE FIXING PROCESS

A hypo solution which is used for fixing must do two things to the emulsion of the film. First, it must stop development and must free the emulsion from all unexposed or unreduced silver salts. Second, it must preserve or tan the emulsion to prevent deterioration. The greater percentage of an emulsion is gelatin and being animal matter it will decompose and deteriorate if not properly treated. The question of how long a film should be fixed is often brought up. A safe rule is to fix twice as long as it takes for the unreduced silver (seen as a yellow creamy substance) to disappear from the film.

LIFE OF DEVELOPING SOLUTION

The developing solution may become exhausted from age rather than from use. However, it is simple to determine the efficiency of the solution by checking the quality of the radiograph developed in it. The first sign of age in a developing solution is discoloration. It will gradually turn yellow and continue to turn dark brown. Developer in such a condition may produce stains and fog and obliterate the contrast that might be otherwise on the film.

LIFE OF THE FIXING SOLUTION

Hypo does not deteriorate rapidly unless films are processed through it. The length of time required to completely fix a film is a reliable check on the strength of the hypo. If it requires more than four times as long to fix a film as when the hypo is new and at the same temperature, it should be replaced.

PREPARATION OF SOLUTION

When mixing new solutions, it is important that the tanks be thoroughly scrubbed. If this is not done, some of the old chemicals will be mixed in with the fresh and the process of oxidation will be set up at once and shorten the life of the new chemicals.

QUESTIONS AND ANSWERS ON DARK ROOM TECHNIC

- Question: What is the most important factor in designing a dark room?
Answer: The dark room must be what the name implies. Absolutely dark. Even a small pinhole of light will effect the film.
- Question: Should the dark room walls be painted black?
Answer: If the room is light proof, the color makes no difference. Black walls are less apt to reflect light that leaks in around doors, etc.
- Question: Can a window in the dark room be painted black to keep out the daylight?
Answer: Paint as a rule is unsatisfactory as it develops pinholes. It is much better to cover the window with some light opaque material.
- Question: What is the most common cause of fogging films?
Answer: Faulty dark room lights. All dark room lights will fog the film if exposed to the light long enough. However, a good dark room light will not fog the film in the ordinary time required to process the film.
- Question: Is the wattage of the bulb used in the dark room light important?
Answer: Very important. Never use a larger lamp bulb than that recommended by the manufacturer. (Usually 15 watts)
- Question: How can the dark room light be tested?
Answer: Take a fresh film without any x-ray exposure on it and cover half of the film with the black paper it is wrapped in. Place the film about three feet from the dark room light and expose it for three minutes. Develop in the regular way and at the correct time for the temperature of the developer. If the light is fogging, the half of the film which was not covered will show fog.
- Question: What kind of developer should be used?
Answer: It is recommended that developer made by the company making the film be used.
- Question: Should regular or rapid developer be used?
Answer: Either regular or rapid developer can be used. However, if rapid developer is used, the speed of the developer can be maintained by the addition of replenisher.
- Question: How often should developer be changed?
Answer: That depends upon how many films are developed and how large the tank is. Developer will exhaust itself without any use in about two or three months. Developer should be changed at least every sixty days.
- Question: Should films be developed by sight or time?
Answer: If the exposure technic is correct, time development is more satisfactory.
- Question: How long should a correctly exposed film be developed?
Answer: That depends upon the temperature. The warmer the developer, the shorter the developing time. Follow the timing directions on the container.
- Question: Should the film be completely drained of developer when removing the film from the developing tank?
Answer: No. The film should be drained for about three seconds. Further draining only drains back into the tank exhausted developer which has been in close contact with the film. Operators should figure on using 1 oz. of developer for each 8 x 10 film, 3-1/2 oz. for each 11 x 17 film, and 7 oz. for each 11 x 36 film.
- Question: How long should the film be washed after removing film from the developer?
Answer: About five or six seconds. The film will keep on developing until it is put into the fixer.
- Question: How long should the film be left in the fixing solution?
Answer: The time is not critical and depends upon the age of the fixer. The film should remain in the fixer at least twice as long as it takes for all the milky appearance of the film to disappear. In warm climates, the film is left in the fixer about 1/2 hour to harden the emulsion.
- Question: How long should the film be washed after removing it from the fixer?
Answer: At least 15 minutes in running water.

- Question: What precautions should be taken in handling the film?
Answer: Hands should be dry and care must be taken not to rub or bend the film. Rubbing the film will put abrasion marks on the film, bending the film will desensitize the film and leave streaks or moon shaped areas.
- Question: What are static marks?
Answer: Black marks on the film which resemble streaks of lightning caused by friction which generates static electricity. This is more prevalent in cold frosty weather.
- Question: How can static marks be minimized?
Answer: Careful handling of the film to eliminate friction. A strip of metal molding on the front edge of the loading bench which in turn is grounded to a water pipe will sometimes carry off the static charge when the cassette touches the molding.
- Question: What causes stains on the film?
Answer: Old worn out solutions or insufficient washing.
- Question: What causes small round white spots on the film?
Answer: Air bubbles. This can be eliminated by raising and lowering the film several times when first put into the developer.
- Question: Should the film be agitated while developing?
Answer: Yes. The film should be raised and lowered in the developer about every thirty seconds. This will bring in fresh developer to replace the exhausted developer which has been in contact with the film.
- Question: Must films be protected from x-ray exposure while in the dark room?
Answer: Yes. Films or loaded cassettes should be stored in a metal protective chest. On some installations a sheet of 1/16 inch sheet lead is fastened to the wall between the film and the source of x-ray radiation. The films must be stored behind the lead when the x-ray equipment is in operation.
- Question: Will it be detrimental if developer gets into the fixer or fixer into the developer?
Answer: Developer in the fixer is not detrimental unless the amount is excessive. However, it only takes a few drops of fixing solution to ruin the developer.

CONCLUSION

Ordinary care and attention to the many small details will improve your radiographic result and make diagnosis much easier.

TROUBLE CHART
DARK ROOM

CONDITION

CAUSE

Gray Films

1. Under developed.
2. Old developer.
3. Developer weakened by dilution.
4. Safe light fog.
5. Developer too cold.
6. Developer too hot.

Stains

1. Improper rinsing.
2. Improper washing.
3. Developer exhausted.
4. Developer contaminated.
5. Exhausted hypo.

Streaks

1. Developer not mixed completely.
2. Under developed.
3. Dirty hanger clips.

Static Marks

1. Rough handling of undeveloped films or cassettes. Friction.

Crescent Marks

1. Bending film while loading.

Frost like Appearance

1. Incomplete washing.

Corner Marks

1. Wet or dirty fingers.
2. Dirty hanger clips.

Blank or opaque areas

1. Films sticking together or to the side of the tank.

Light Spots

1. Fixer or water on film before development.

Dark Spots

1. Dust, liquid or powdered developer on film before development.

White Spots

1. Bubbles clinging to emulsion surface during development.
2. Dry hypo on wet film while drying.

Lack of Contrast

1. Under developed.
2. Cold developer.
3. Exhausted developer.

Multiple White Specks

1. Dirty Screens.
2. Barium on cassettes, sheets or table top.

IMPROVING THE RADIOGRAPH

Question: How can I improve detail or sharpness of the radiographic image?

Answer: Do everything possible to stop motion. Compression bands as a rule are not desirable as they cause distortion of the spine. However, a head clamp will eliminate a great deal of motion without distorting the spine. Another factor that will improve detail is to increase the focal distance. This calls for an increase in the time factor. Refer to the change of distance table on all Fischer charts to determine how much time increase is required for any desired distance. The figures shown on the change of distance table are multiplying factors and should be used to multiply the time factor shown at the standard distance. (For example) If the regular technic on a pelvis calls for 25 MA, 5 seconds at 30 inches and it is desired to increase the distance to 36 inches to improve detail, find 30 inches in the left hand column of the change of distance table then read straight across to the 36 inch vertical column. Where the 30 inch horizontal line intersects the 36 inch vertical column will be found the figures 1.4. This is the multiplying factor required to obtain the same density at 36 inches that was obtained at 30 inches. Multiplying the time factor of 5 seconds for 30 inches by 1.4 gives us a time factor of 7 seconds for 36 inches. This added time will compensate for the increased distance and produce better detail than that obtained at 30 inches. Care must be taken to make sure that the increased time will not cause motion on the part of the patient.

Question: How can radiographic contrast be improved?

Answer: By contrast we mean the percentage difference between the blacks and whites on the film.

There are many factors that effect contrast. Unsafe dark room lights, light fog, x-ray fog, outdated film, worn out solutions, etc. (See questions and answers on dark room technic). Another factor that effects contrast is secondary radiation. Secondary radiation is created in the tissues and gives the appearance of a fog. The thicker the mass and the denser the mass being x-rayed, the greater the amount of secondary radiation generated. Secondary radiation can be reduced by using a cone that just covers the area that the operator wishes to show on the film. It is obvious that reducing the size of the port of entry by the use of a cone, that secondary radiation will not be generated in tissues surrounding the anatomical area under examination.

Another factor which effects secondary radiation is kilovoltage. The higher the voltage the more secondary radiation generated in the tissues. Thick masses of tissue usually call for higher kilovoltage. If the kilovoltage is reduced and the time increased to compensate for the reduced kilovoltage, much better contrast will be obtained on heavy patients. Referring to the relation of KVP to time of exposure table on all Fischer technic charts, lets assume that we want to reduce the kilovoltage 5 KVP to improve contrast. Under the column with the heading (Reduce KVP) find the figure 5. On the same line in the next column to the left under the heading (Increase Time) will be found the figures 1-1/2. This is the multiplying factor to be used to compensate for the reduction in voltage. For example - If the chart calls for 25 MA, 5 seconds and 70 KVP and we decide to reduce the KVP 5 KVP, it will be necessary to multiply the time factor of 5 seconds by 1-1/2 which gives us 7-1/2 seconds time. If we were to reduce the KVP 10 KVP according to the table, we would have to multiply the regular time factor by 2 which

would make our exposure 10 seconds instead of 5. If for any reason it was desired to increase the KVP in order to reduce time, the two columns on the left hand side of the table can be used. It will be noted that increasing the KVP 5 KVP will reduce the time to $3/4$. If we increase the KVP 10 KVP the time can be cut in half.

Question: Is it necessary to make corrections in exposure time for patients of different ages?

Answer: Yes, in some cases. No change is necessary on patients between 14 and 55 years of age. Referring to the age correction table, it will be noted that on patients over 55 years of age, the KVP should be reduced 2 KVP or the exposure reduced to 80 per cent of normal. The table also shows the reduction necessary in patients of from 1 to 12 years of age. Use of this table will permit more uniform results.

SUMMARY

Better radiographs will be obtained by:

- 1- Making sure there is no motion on the part of the patient.
- 2- Making sure all your dark room factors are correct.
- 3- Reducing kilovoltage and increasing distance where greater time factor is practical.

Initial Anode-Film Distance	TO CHANGE FOCAL DISTANCE FACTORS							
20"	1.0	1.6	2.3	3.2	4.0	5.8	9.0	13.0
25"	.64	1.0	1.4	2.1	2.6	3.7	5.8	8.3
30"	.44	.69	1.0	1.4	1.8	2.6	4.0	5.8
36"	.31	.48	.69	1.0	1.2	1.8	2.8	4.0
40"	.25	.39	.56	.81	1.0	1.4	2.3	3.2
48"	.17	.27	.39	.59	.69	1.0	1.6	2.3
60"	.11	.17	.25	.36	.44	.64	1.0	1.4
72"	.08	.12	.17	.25	.31	.44	.69	1.0
New Anode-Film Distance	20"	25"	30"	36"	40"	48"	60"	72"

RELATION OF KVP TO TIME OF EXPOSURE			
REDUCE TIME	ADD KVP.	INCREASE TIME	REDUCE KVP.
3/4	5	1-1/2	5
2/3	8	2	10
1/2	10	2-1/2	12
TO 3/5	12	TO 3	13
1/3	13	4	16
1/4	16	4-1/2	17
1/5	18	5	18

AGE CORRECTION		
BY EXPOSURE OR KILO-VOLTS		
PER-CENT EXPOSURE	AGE	REDUCE K.V.P.
100%	14 to 55	0 K.V.
90%	12 Years	1 K.V.
82%	10 Years	2 K.V.
72%	8 Years	3 K.V.
64%	6 Years	4 K.V.
56%	4 Years	6 K.V.
48%	2 Years	8 K.V.
44%	1 Year	9 K.V.
Over 55 Years Employ 80% or -2 K.V. less.		

For an explanation of the above tables refer to article
on improving the Radiograph.

QUESTIONS AND ANSWERS ON THE USE AND CARE OF INTENSIFYING SCREENS AND CASSETTES

- Question: What are intensifying screens?
Answer: Intensifying screens are made with a cardboard base and coated with a fluorescent material which emits visible light when excited by x-ray radiation.
- Question: How are intensifying screens used in radiographic work?
Answer: The intensifying screens are mounted in a light proof holder called a cassette. The film is loaded in the cassette in a dark room. Both fluorescent surfaces of the screens are in direct contact with the film.
- Question: Just what happens when x-radiation strikes the screens?
Answer: The screens fluoresce with visible light in direct proportion to the amount of x-ray radiation striking the screens. The result is that the film is not only exposed with x-ray but is exposed with visible light.
- Question: What is the advantage of this?
Answer: It permits a much shorter exposure time.
- Question: What is the intensifying factor when using intensifying screens?
Answer: The intensifying factor will vary with the kilovoltage. On an average it will be about 40 to 1.
- Question: What percentage of the exposure is made with x-ray and what percentage is made with visible light when using intensifying screens?
Answer: About 98 percent of the exposure is made with visible light and 2 percent with x-ray radiation.
- Question: What advantage does the shorter exposure time give?
Answer: Eliminates movement by the patient and saves wear and tear on the x-ray tube and machine.
- Question: What films should be used with intensifying screens?
Answer: Regular films.
- Question: Why?
Answer: Regular films are very sensitive to the blue light emitted from the intensifying screens.
- Question: Can non screen films be used with intensifying screens?
Answer: Not satisfactorily because non screen films are not sensitive to visible light.
- Question: Are all screens the same speed?
Answer: No. Intensifying screens come in three different speeds. Slow or definition, medium speed and extra fast.
- Question: What determines the speed of a screen?
Answer: The size of the crystals used. The larger the crystals the faster the screen.
- Question: Do high speed screens give as good detail as slower screens?
Answer: No. The slower the screen, the better the detail. However, high speed screens as manufactured today give very satisfactory radiographs.
- Question: What determines the selection of screens?
Answer: The power of the equipment. Generally high speed screens are used on equipment up to 100 milliamperere capacity. Par-speed or medium speed screens are used on 200 or 250 milliamperere machines except on full spine work taken at a 60 inch focal distance in which case, high speed screens should be used.
- Question: What care should be taken with screens?
Answer: Screens should be inspected every week and all dust particles cleaned out with a camel hair brush. If screens are soiled, clean with a wad of cotton soaked in non-medicated grain alcohol or carbon tetrachloride. Developer stains cannot be removed. Cassette should always be closed when not in use.
- Question: If the screens turn yellow will it effect the speed of the screen?
Answer: Yes. Screens should be discarded if they show a yellow tinge.

Question: How should a film be removed from the cassette?
Answer: After releasing the spring bars, turn cassette upside down and lift bakelite side of cassette up. Film will drop out. This eliminates digging the film out with the finger nails and possibly ruining the screen.

Question: What care should be given cassettes?
Answer: Cassettes should be handled carefully. Dropping the cassette on the floor will very likely spring the cassette and impair the contact. Examine the bakelite backs from time to time and note if there is any bulging of the bakelite. If bulging is apparent the chances are that poor screen contact is present with the consequent loss of detail.

QUESTIONS AND ANSWERS ON BUCKY DIAPHRAGMS AND STATIONARY GRIDS

- Question - What is the purpose of a bucky diaphragm or stationary grid?
 Answer - The purpose of a bucky diaphragm or stationary grid is to filter out secondary or scattered radiation which is generated in the tissues and tends to fog the radiograph.
- Question - How does the filtering process function?
 Answer - Both the bucky diaphragm and stationary grid, incorporate a grid which is made up of alternate strips of lead and wood set on edge. X-radiation traveling in straight lines will go through the grid while scattered or secondary radiation will strike the lead strips and be absorbed.
- Question - What is the difference between a bucky diaphragm and a stationary grid?
 Answer - A bucky diaphragm is equipped with a motor mechanism which moves the grid continuously across the film while the exposure is being made. This movement eliminates the shadow of the lead strips. In the case of a stationary grid, the grid does not move so consequently the lead strips show on all films.
- Question - Is there any advantage in using a stationary grid instead of a bucky diaphragm?
 Answer - None except lower cost and portability for bedside work.
- Question - What determines the efficiency of a bucky or stationary grid?
 Answer - The filter ratio. The higher the ratio the more secondary radiation filtered out.
- Question - What determines the filter ratio?
 Answer - The ratio between the thickness of the grid and distance between the lead strips. If the thickness of the grid is six times the distance between the lead strips it would have a 6 to 1 ratio.
- Question - What is meant by a 50 or 16 line grid?
 Answer - A 50 line grid means that there are 50 lead strips to the inch. A 16 line grid has 16 lead strips to the inch.
- Question - Does a 50 line grid have any advantages over a 16 line grid?
 Answer - Yes. A 50 line grid can be made much thinner and still maintain the same filter ratio. The thinner grid has less absorption.
- Question - How should the bucky diaphragm be timed in relation to the x-ray exposure?
 Answer - The timing on the bucky should be about 20% longer than the actual x-ray exposure.
- Question - Are the timing scales on a bucky accurate?
 Answer - The timing scales are usually quite accurate when the room temperature is around 70°. At higher temperatures they will be fast and at lower temperatures they will run slow. This is due to a change in the viscosity of the oil used in the cushion chamber which will vary with the temperature.
- Question - What causes grid lines on the film when using a bucky?
 Answer - Several things can cause grid lines.
 1st. Uneven travel of the grid due to air in the oil chambers, low oil, sticky bearings or binding of the grid or grid mechanism or some adjacent part.
 2nd. Ratio of grid travel to actual timing too great.
 3rd. Grid timing too short. Grid timing should be about 20% greater than actual x-ray exposure.
 4th. Tube not centered on the bucky.
 5th. Too great or too little tube to grid distance. (focal distance)
- Question - Is it possible for all technical factors to be correct and the bucky functioning perfectly and still show grid lines?
 Answer - Yes. On self-rectified units certain grid timing and x-ray timer combinations

will create a stroboscopic effect which will show grid lines even though all factors are perfect. The remedy is to change the timing of the bucky slightly.

Question - What is meant by focal distance on a bucky diaphragm or stationary grid?

Answer - The grids on a bucky and an aligned stationary grid are so designed that the lead strips are radial to focal spot of the x-ray tube at a specific tube to film distance.

Question - What focal distances are buckys built for?

Answer - Bucky diaphragms can be obtained in focal distances of 30 inches, 36 inches and 48 inches. Aligned stationary grids can be obtained in 30 and 36 inch focal distances.

Question - If a bucky or stationary grid is built for a certain focal distance does that mean that no other distance can be used?

Answer - No. A 30 inch bucky can be used at focal distances of from 25 inches to 40 inches. A 36 inch bucky can be used from 30 inches to 45 inches and a 48 inch bucky from 45 to 60 inches.

Question - Is there any difference in exposure time required if the focal distance is more or less than that which the bucky or grid was designed for?

Answer - Yes. A greater or shorter focal distance will increase the absorption in the grid and require a longer exposure time.

Question - Can a bucky designed for 48 inch focal distance be used at a 30 inch distance?

Answer - No. A 48 inch bucky cannot be used at less than 45 inches with satisfactory results.

Question - Are all stationary grids aligned?

Answer - No. Stationary grids come in both aligned and paralleled types.

Question - Has the aligned grid any advantage over the parallel grid?

Answer - Yes. Due to the fact that the aligned grid is focused for a specific focal distance it is more efficient.

Question - When is the parallel grid used?

Answer - For long distance work such as chest taken at 72 inches.

STEREOSCOPIC WORK
ON ALL T UNITS, MPA20 & MPA30
X-RAY UNITS

From time to time users and prospective customers of T Units, MPA20 and MPA30 Unit express the desire to make stereoscopic films. To make stereoscopic films of any anatomical region, it is necessary to make two films from different angles. The tube being shifted an equal distance each side of center to obtain the necessary angulation. The films must be changed without disturbing the patient which calls for the use of a cassette tunnel or Bucky diaphragm.

Due to the fact that the cone port on the T Units, MPA20 or MPA30 is 1-1/2" off center of the head, a 3" stereoscopic lateral shift can be obtained by manipulating the head in a fashion that will permit making two radiographs at different angles. The following sketches along with a step by step description of the head manipulation will allow the operator to make the necessary stereoscopic shift quickly and easily.

- No. 1 Center head (not cone port) on center of Bucky or film tunnel. The Allen set screw holding the tube head in the yoke can be used as a centering point. Centering the head as described automatically placed the cone port 1-1/2" off center (see Fig. 1). This is the position for the 1st stereoscopic film.
- No. 2 After changing films without disturbing the patient, rotate tube head in the yoke 180° or so that tube port points up or away from the table. (See Fig. 2)
- No. 3 Rotate tube on its horizontal Axis or cross arm 180° so that cone port points down or towards the table. This automatically places cone port 1-1/2" the other side of center, making a total shift of 3". (See Fig. 3).

Fig. - 1

Position of head
described in Step
1.




Fig. - 2








Fig. - 3



When the T Unit is mounted on rails such as is the case when using a vertical Bucky a stereoscopic shift can be made by following the step by step instructions below:

- No. 1 Mount rails and Bucky so cone port centers on Bucky when head is in the position shown in Fig. 4. With the tube head in this position non-stereoscopic films can be made with the tube centered on the Bucky.
- No. 2 Make 1st film by rotating tube head 90° on horizontal axis or cross arm, placing head in position shown in Fig. 5. This places cone port 1-1/2" to one side of center.

No. 3 After changing films, rotate tube head 180° on horizontal axis or cross arm placing tube head in position shown in Fig. 6. This places tube head 1-1/2" the other side of center, making a total shift of 3".



While it is customary to shift the tube 10% of the focal distance, this shift is not critical. The 3" shift obtained on the TC, MP 10 and 20 can be used on any focal distance of from 25 to 42".

It will be found that with a little practice, a stereoscopic shift as described above can be made very quickly and with ease.

RULES FOR STEREOSCOPIC FILMS

- 1 - Patients position must not be disturbed during the entire process of making the two films. Use a cassette tunnel or Bucky diaphragm which will permit changing films without moving the patient.
- 2 - Place second film in exactly the same position as the first film.
- 3 - It is not necessary to tilt tube head to focus on the center of the film unless cone used is small enough to cause cut off.
- 4 - Stereoscopic shift must always be made across the long bones of the body.
- 5 - The stereoscopic shift on chest should always be made across the ribs.
- 6 - Stereoscopic films on chest made with a manual shift and manual changing of the films is very unsatisfactory as both films must be taken on a single inhalation. Stereoscopic chest work calls for an automatic stereo shift and an automatic cassette shifter.
- 7 - Stereoscopic films should have the same density.
- 8 - In order to see stereoscopically, it is necessary to have eyesight in both eyes. Some people with good eyesight cannot see stereoscopically.
- 9 - Films must be placed in stereoscope properly in order to obtain the proper perception.

NOTE:

The above method and procedure applies to any self contained head having an off set cone port.

BULLETIN

X-RAY PROTECTION AT THE TUBE HEAD

We are all aware of the dangers of x-ray radiation both to the operator and the patient. With modern day equipment, screens and films, the amount of exposure received by the patient is relatively small compared to the total amount of radiation that can be tolerated. The chief danger from the standpoint of the patient is repeated exposures or prolonged fluoroscopic examinations. However, x-ray exposure to the operator is a different story.

Due to the fact that x-ray exposure is accumulative and a tolerance is never established, such as in the case of sunburn, it is vitally essential that every precaution be taken to protect the operator who is exposed to small amounts of radiation day after day, which cannot be measured.

From the foregoing it is obvious that every effort must be made to eliminate all stray radiation from the tubehead and confine the x-ray beam so that only the radiation used in making the radiograph or fluoroscopic examination emanates from the tubehead.

All x-ray manufacturers attempt to build into their tubeheads some sort of x-ray protection. The method many competitors use, is to build into their equipment just enough protection to get by, or in other words, what they think is good enough. The other method, THE FISCHER WAY, is to build into every tubehead x-ray protection in accordance with the Bureau of standards specifications.

This is an unseen and costly feature on FISCHER-built x-ray apparatus and the doctor or hospital can only become aware of this added value by having the FISCHER representative explain in detail the effort and expense the Company puts forth to give the operator adequate x-ray protection.

DESCRIPTION AND USE OF THE FISCHER ALIGNING CONE

In many types of x-ray work it is essential that the central ray emanating from the x-ray tube be directed accurately through certain anatomical parts in order to show the part clearly on the film without distortion or super-imposition. This is particularly true where it is necessary to angulate the x-ray tube, such as in cervical spine and sinus work.

In the past, aligning the central ray of the tube with the anatomical part has been done largely by guess work, using the eye to line up the central ray with the part. This was very inaccurate and the results very undependable.

Using the aligning cone is a very simple procedure where the machine is operated on properly placed tracks. The hole in the base of the cone in which the cord is attached should be on the operating side of the head, 90° from the perpendicular. To trace the course of the central ray from the source at the x-ray tube to the spot on the film where it will terminate, it is only necessary to line up the cord attached to the cone base with the aligning indicator on the cone proper and then stretch the cord to the spot on the bucky where the central ray will strike. Angulating the tube, adjusting the tube height and bucky height as well as angulating the patient's head is then a simple procedure. The aligning cord will also show just what anatomical structures the central ray will transverse.

In addition to the above, rotating the cone 90° so that the aligning cord and indicator are on top of the cone will permit accurate centering of the tube on the bucky in installations where mobile units not on tracks are used.

The above item is cataloged as follows:

Aligning Cone for use with any x-ray machine presently manufactured by us. Same as Cone "C" (Cat. No. 4736) except that it has aligning feature. When used with the Model TF our Cat. No. 4742 Adapter must be ordered additionally.

H. G. FISCHER & CO.

THE IMPORTANCE OF ADEQUATE
POWER SUPPLY FOR X-RAY
INSTALLATIONS

Occasionally we receive letters of complaint from doctors stating that their X-Ray installation is not producing the quality of radiographic work expected or desired.

There are many reasons why satisfactory results are not obtained, such as poor dark room technic, slow films, worn out screens, ect. This phase of the situation will be covered in another bulletin.

We have found that in many cases, the difficulty is due to inadequate power supply. If the line wires are too small for the load or the pole transformer is not large enough, there will be an excessive drop in voltage when the x-ray load is thrown on the line. This in turn causes a drop in KVP on the tube. The doctor might have the controls set for a given KVP (for instance) 85 KVP - 100 MA. However, due to excessive line drop he is perhaps only obtaining 75 KVP when the x-ray load is thrown on the line. This naturally causes a light film which can only be corrected by adding time or KVP.

Another factor that enters into the picture when excessive line drop is present, is that the inverse voltage increases tremendously causing added stress on cables, tube and transformer insulation. Many of the punctured cables are no doubt due to high inverse voltage caused by excessive line drop.

The attached wiring table indicates the size wire and transformer capacity required for all Fischer machines necessary to obtain satisfactory operation. The wire and transformer sizes indicated, are figured to give a line drop of not more than 3%.

NOTE: IF WIRE SIZE INDICATED IS NOT AVAILABLE, USE NEXT LARGER SIZE.

WIRING SPECIFICATIONS

TC 30 - TF 30 - DELUXE TC 30 Operating on 220V - 60 Cy

DISTANCE FEET - 50'	75'	100'	150'	200'	250'	300'	350'	400'
WIRE SIZE NO. - 12	10	10	8	8	6	6	4	4
CUT OUT BOX - 30 AMP.								
FUSES - 30 AMP.								
AMPERES - 25 AMP.								

TC 30 - TF 30 - DELUXE TC 30 Operating on 110V - 60 Cy

DISTANCE FEET - 50'	75'	100'	150'	200'	250'	300'	350'	400'
WIRE SIZE NO. - 8	6	4	2	2	1	1/0	2/0	2/0
CUT OUT BOX - 60 AMP.								
FUSES - 60 AMP.								
AMPERES - 50 AMP.								

TC 50 - TF 50 - DELUXE TC 50 Operating on 220V - 60 Cy ONLY

DISTANCE FEET - 50'	75'	100'	150'	200'	250'	300'	350'	400'
WIRE SIZE NO. - 10	8	8	6	4	4	2	2	2
CUT OUT BOX - 60 AMP.								
FUSES - 60 AMP.								
AMPERES - 50 AMP.								

DELUXE RF-100 and RF-100 Operating on 220V - 60 Cy ONLY

DISTANCE FEET - 50'	75'	100'	150'	200'	250'	300'	350'	400'
WIRE SIZE NO. - 8	6	6	4	2	2	1	1	1/0
CUT OUT BOX - 100 AMP.								
FUSES - 75 AMP.								
AMPERES - 75 AMP.								

DELUXE RF-250 Operating on 220V - 60 Cy. ONLY

DISTANCE FEET - 50'	75'	100'	150'	200'	250'	300'	350'	400'
WIRE SIZE NO. - 8	6	4	2	2	1	1/0	2/0	2/0
CUT OUT BOX - 100 AMP.								
FUSES - 100 AMP.								
AMPERES - 100 AMP.								

February 23, 1950

MILLIAMPERE PRE-ADJUSTMENT SWITCH ASSEMBLY

Experience has proven that more X-Ray tubes are damaged in testing the tube for proper milliamperage values than in actual radiographic exposure. This is particularly true when using high milliamperage values, such as used in chest or gastrointestinal work where high milliamperage values and short exposure time is necessary. In the case of a chest radiograph, the time of exposure is only 1/10 or possibly 2/10 of a second. However, if the control is adjusted to the proper milliamperage value by actually passing high tension current through the tube, a minimum of 1 second is consumed in testing, as it takes that long for the milliamperage meter to register the correct value. The chances are that this one-second test exposure will be made two or three times before the milliamperage factor is adjusted correctly.

All time consumed in testing the X-Ray tube for correct milliamperage must be computed in the loading limit of the tube. It is easily possible for an operator to load the X-Ray tube to the limit just testing and then overload the tube when making a very short exposure, although the actual radiographic exposure is well within the load limit of the tube. So much time has been used in testing that the short radiographic exposure used is enough to exceed the heat storage limit of the X-Ray tube and tube damage results.

is a modern X-Ray apparatus, when equipped with the milliamperage pre-adjustment switch, eliminates the necessity of passing high voltage through the X-Ray tube when testing the tube for correct milliamperage values, thus saving wear and tear on the X-Ray tube with resulting tube economy.

To use the milliamperage pre-adjustment switch, it is necessary to calibrate the unit at the time of installation. Assuming that the unit being calibrated is a 50 MA TC, which has kilovoltage scales for 10, 30, 40 and 50 MA, all four milliamperage values are to be calibrated. On the blank chart furnished with all machines with pre-adjustment switches, (see sample below) enter in the blank squares on the milliamperage line of the chart the milliamperage values to be calibrated, which in the case of the TC 50 will be 10, 30, 40 and 50.

MILLIAMPERE PRE-ADJUSTMENT FOR

"T" Models 50-96

MILLIAMPERES

PRE-ADJUSTMENT
READING

- 1st. Adjust the kilovoltage for approximately 50 KV.
- 2nd. With the timer switch on the test position, depress the timer push button and adjust the main and auxiliary milliamperage controls until milliamperage meter reads 10 MA.
- 3rd. Release timer push button and depress pre-adjustment switch located on top of control. Note meter reading on red scale which is the top scale on the meter. Record this reading on the calibration chart under 10 MA and on the pre-adjustment line.
- 4th. Repeat for 30, 40 and 50 milliamperes and record the readings obtained on the red scale under the proper milliamperage value on the calibration chart.

NOTE: The red pre-adjustment scale on the kilovolt meter is not calibrated in milliamperes but is an arbitrary scale and bears no direct relation to the actual milliamperage values. Upon calibrating the unit, it might be found that 30 milliamperes are obtained when the meter reads 22 on the pre-adjustment scale. At any future time 30 milliamperes can be obtained by depressing the pre-adjustment switch and adjusting the main and auxiliary milliamperage controls until the meter reads 22 on the red pre-adjustment scale. No actual high tension current is passing through the tube.

It is obvious from the foregoing that there is a distinct advantage, when using high milliamperage, in being able to pre-adjust the milliamperage values without actually passing high voltage through the tube. With the idea in mind that greater tube economy will be forthcoming if a pre-adjustment device is used, it has been decided to incorporate the pre-adjustment device in all 50 milliamperage units as standard equipment.

The following x-ray units of H. G. Fischer & Co. manufacture will have incorporated in them, without extra charge, a Pre-adjustment Switch Assembly:

Cat. No.

5137R	TCR-50 Milliamperage X-ray Unit on Rails
5137R-L	Same as Cat. 5137R but with longer upright
5138	Deluxe Rail Mounted TC-50 Milliamperage X-ray Unit
5136	TC-50 Milliamperage Mobile X-ray Unit
5137	TC-50 Milliamperage Rail Mounted X-ray Unit
5149	TF-50 Milliamperage X-ray Unit
5185	Spacesaver 50 Radiographic-Fluoroscopic Unit and Examining Table

The Fischer Model RF-100 Milliamperage Shockproof X-ray Apparatus with Standard Controls and the Fischer Model L-200 X-ray Apparatus incorporate a Pre-adjustment Switch that differs somewhat from the Pre-adjustment arrangement on our "T" Units.

The Fischer De Luxe RF-100 Milliamperage X-ray Apparatus and the Fischer De Luxe RF-200 Milliamperage X-ray Apparatus do not require a Pre-adjustment arrangement since the automatic controls of these units eliminate the necessity for this device.

GENERAL INFORMATION

PATIENT COMFORT

Every effort should be made to make the patient as comfortable as possible. All equipment adjustments possible should be made before the patient is placed on the table. The longer the patient is on the table the greater the chances are for motion. When radiographing extremities, the use of sandbags will help in eliminating motion.

VARIATIONS IN TECHNIC

All technic charts are made up for average patients. Muscular patients require 20% to 30% more exposure. Soft, flabby patients require about 20% less exposure.

MEASURING THE PATIENT

The patient should always be measured in the position that the radiograph is to be taken. If the patient is to be radiographed in the horizontal position, that is the position in which the patient should be measured. If the patient is to be radiographed in the standing position, the patient should be measured in the standing position. Always make sure that the caliper just lightly touches the patient on both sides and measurement is made at the thickest part of the area to be radiographed.

AGE CORRECTION

Patients between the ages of 14 and 55 years normally have tissues of the same densities and standard technic can be employed. Patients over 55 years of age usually have a lower bone calcium and for that reason the kilovoltage should be reduced 2 kilovolts or the time reduced 20%. Patients between the ages of 1 and 14 years require less exposure, depending upon the age. By referring to the age correction table on the technic chart, the necessary correction in exposure or kilovoltage can be made for any age from 1 to 14 years.

THE USE OF CONES

Few operators realize the importance of cones when doing radiographic work. The cone used for any radiograph should be just large enough to cover the anatomical area under examination. Using a cone just large enough to cover the desired area will reduce secondary radiation from surrounding tissues and produce a radiograph with much more contrast. Most charts furnished with x-ray units indicate what cone to use for each anatomical part.

FLUOROSCOPY

Before attempting to do fluoroscopic work, carefully read the section on fluoroscopy. Care should be exercised to make sure that the fluoroscopic shutters are in the open position when doing radiographic work. An arrow on the fluoroscopic shutter control casting indicates the direction the shutters are to be moved for the open position.

DARK ROOM

90% of radiographic failures are due to dark room difficulties. Careful study of the section on dark room technic will reduce radiographic failures to a minimum.

NON BUCKY TECHNIC

Should it be desired to dispense with the use of a bucky diaphragm or stationary grid and use only a cassette equipped with hi-speed screens, comparative densities can be obtained by reducing the kilovoltage 13 kilovolts below that shown for the thickness of the part or use 1/3 of the time indicated when using a bucky diaphragm. Better results will be obtained if the kilovoltage is reduced and the time factor remains the same as indicated for bucky technic.

POSITIONING INFORMATION

Depth measurement - The measurement in centimeters for any particular structure should be along the line of central ray projection.

Chest AP - Measure at plane between 5th and 6th dorsal and approximately 3 to 5 c.m. one side of the spine medium line. Chest should be taken in the standing position using a wall mounted cassette holder. Place back of hands on patient's hips. Bend elbows out to permit shoulders to come as close to the film as possible. Center tube at 5th or 6th dorsal.

Chest LATERAL - Measure patient with arms down, sliding centimeter caliper between arms and patient's body. Make measurement on a line with lower margin of scapula. Patient's arms are then raised above head. Center tube on line of measurement.

A.P. Lumbar-Pelvis - Caliper at point $2\frac{1}{2}$ " below sternum as a measure of both full kidney and urinary tract and full lumbar and pelvis. For spine and bone work, the entire portion of the body between a point $2\frac{1}{2}$ " below sternum through the first two-thirds of the femur may be considered lumbar pelvic grouping. The lower third of the femur takes the extremity grouping. Center tube at 3rd or 4th lumbar.

Lumbar and Sacrum LATERAL - For lateral lumbar spine center spine with center of table. Measure $1\frac{1}{2}$ " above crest of ilium and center tube at that point.

Cervical Spine A.P. and LATERAL - Measure in plane of cervical dorsal junction.

LATERAL Sinus - Measure across forehead approximately on line drawn between frontozygomatic sutures and center one-half way along line between external meatus and outer canthus.

Frontal Sinus P.A. - Place patient face down on table. Have nose and forehead touching table top. Tilt x-ray tube 12° towards the feet. Direct central ray to center of film.

Maxillary Sinus - Patient in prone position. With full weight of head resting on the chin, the head is tilted until a line between the external meatus and the outer canthus of the eye forms an angle of 37° with the table top. X-ray tube is not angled. Direct central ray to center of film.

Sphenoid Sinus, Open Mouth - Patient in prone position. Head resting on chin and nose with mouth open. Tilt tube 20° towards patient's feet. Central ray should pass half way between upper and lower lips when mouth is open.

Mastoid LATERAL - Patient is placed in prone position. Permit head to assume a natural position, with the weight on the cheek. Tilt tube 15° toward patient's feet. Center $1\frac{1}{2}$ " above and $1\frac{1}{2}$ " behind the tip of the ear. When possible, fold ear forward.

Sternum Oblique - Place patient in prone position. Rotate patient slightly to eliminate superimposition of spine over sternum.

Atlas and Axis - Patient is placed on the table in the supine position with the median line of the body over the center line of the table. The chin is elevated until the occlusal surface of the teeth is vertical with the table top. With the mouth wide open, the film is entered half way between upper and lower lip. The central ray is directed to the center of the film.

Wrist P.A. - The forearm is placed on the table top with the palmar surface down. The wrist is centered to the film by placing the midpoint between the styloid processes over the center of the film. Direct central ray to center of film.

Wrist LATERAL - Place styloid process of the ulna over center of the film. Direct central ray to center of film.

Elbow LATERAL - When possible the arm is flexed to approximately 45° angle. The hand is turned sideways with thumb up. The medial epicondyle is centered on the film. Direct central ray to center of the film.

Elbow A.P. - The back of the arm and forearm are placed on film holder. The midpoint between the epicondyles is placed in the center of the film. Direct central ray to film. Place a sandbag in patient's hand.

Shoulder A.P. - Patient in supine position. Elevate opposite shoulder with sandbags or pillow to bring shoulder to be radiographed closer to film. Arm is abducted with palmar surface of hand upward. Place sandbag in patient's hand.

Foot A.P. - Patient may lie or sit on table with knee flexed at an acute angle, the plantar surface of the foot resting on film holder. The proximal end of the second metatarsal is centered to the film. Tilt tube 15° towards patient's head and direct central ray to center of film.


Foot LATERAL - Patient is placed in the lateral position with leg slightly flexed and supported by sandbags at knee. A point approximately $1/2$ the distance from the great toe to the heel is centered on the film. Direct central ray to center of film.

Ankle A.P. - Place patient in the supine position with heel resting on film holder. Place sandbag under knee. The lower tip of the internal malleolus is centered on the film. Direct central ray to center of film.


Ankle LATERAL - Place patient in lateral position. The ankle is placed on the film with the external malleolus on the film. The opposite leg is crossed over. The lower tip of the external malleolus is centered on the film. Direct central ray to center of film.

Knee LATERAL - Place patient on table with affected knee resting on film. Center lower margin of patella on center of film. Direct central ray to center of film.


Knee P.A. - Place patient in prone position. Place sandbags under ankles to support toes. Lower margin of patella is placed in center of film. Direct central ray to center of film.




SINUS LAT.




MASTOID FOR BUCKY



MASTOID DOUBLE ANGLE




SINUS




FRONTAL 23°-SPHENOID 17°



JAW



GI. PRONE




GI. ERECT




PA. CHEST




LAT. CHEST



LAT. CERVICAL




AP. CERVICAL




LAT. LUMBAR





AP. LUMBAR




LAT. LUMBAR




AP. DORSAL





HAND PA.




WRIST LAT.




ELBOW LAT.




ELBOW AP.




KNEE AP.




KNEE LAT.



FOOT AP.



FOOT LAT.



OS-CALCIS

FACTORS TO BE CONSIDERED WHEN PURCHASING X-RAY EQUIPMENT

Question: What are the chief factors that should be considered when purchasing x-ray equipment?

Answer: There are several factors that should be considered when anticipating the purchase of x-ray equipment.

1. The integrity and reputation of the company should be considered.
2. The ability to service equipment in the field is very important.
3. Is the equipment of modern design?
4. Does the equipment offered have adequate power for the work you intend to do?
5. Does the unit have a fine and flexible control?
6. Is the equipment especially designed for your purpose or is it something designed years ago and adapted to your particular specialty?
7. Is the price in keeping with the design and output of the equipment?

Question: What bearing has the integrity and reputation of the manufacturer have on purchasing x-ray equipment?

Answer: The integrity and reputation of the manufacturer is of the highest importance. H. G. Fischer & Co. has established a reputation over a long period of years for fair dealing with the profession and pricing of equipment in keeping with the quality offered. The origin of H. G. Fischer & Co. dates back 39 years and is the oldest manufacturer of x-ray and physical therapy equipment operating under the original company name and management.

Question: Why is field service so important?

Answer: All electrical and mechanical devices require occasional adjustment. H. G. Fischer & Co. maintains a nation wide field service organization for the purpose of rendering efficient service on your equipment. All Fischer representatives are factory trained.

Question: What is meant by modern design?

Answer: By modern design we mean equipment that is especially designed for a particular use and incorporating all the modern features such as automatic selection or presetting of milliamperage, flexible kilovoltage controls, stereoscopic shift and other features which go to make technic a matter of precision.

Question: What is meant by adequate power?

Answer: The power required depends upon the time factor in seconds that the operator desires to use. The permissible time used is dependent upon the ability of the patient to remain motionless. Heart and chest films should be taken at 72 inches and 1/10 second to eliminate heart motion and reduce distortion. The stomach usually must be taken in 1/2 second to eliminate peristaltic action. Full spines are much more satisfactory when made at 60 inches with high milliamperage values as the higher milliamperage permits shorter exposure time with less chance of motion.

Special attention should be paid to the tube transformer ratings.. Reputable manufacturers do not hesitate to show the kilovoltage obtained at any given milliamperage load. Some manufacturers will rate a unit as 100 KV - 100 MA. However, a check will show that 100 milliamperes can only be obtained when using a low kilovoltage value, such as 65 to 75 kilovolts. This situation is especially prevalent on so called 30 milliamperage - 90 kilovolt machines which in many cases produce only about 60 kilovolts at 30 milliamperes. The 90 kilovolts is obtained only when using 5 milliamperes.

Question: Why should the x-ray control be fine and flexible?

Answer: Modern day x-ray technic calls for a centimeter thickness of part technic. The patient is measured with a centimeter caliper and the kilovoltage is

adjusted for that particular thickness of tissue. As each centimeter of tissue requires a change of 2 kilovolts, it is obvious that x-ray controls with 9 to 15 steps of control are not adequate as each step of voltage control represents kilovoltage jumps of 5 to 8 kilovolts. With coarse voltage steps as described, it is impossible to accurately adjust the kilovoltage for the thickness of part. The only alternative is to compensate with time for the discrepancy in voltage which necessitates that the operator resort to hunch or guesswork. It requires years of experience to become a good hunch technician.

Question: What is meant by specialized equipment?

Answer: Many manufacturers have one standard design and try, by making a few minor changes to adapt this design to all types of x-ray work. This has proven to be very unsatisfactory. H.G. Fischer & Co. have always built special equipment for the particular specialty. In designing equipment for any special radiographic work, every attempt is made to consult with outstanding men in the profession in order to give the user the ultimate in efficiency.

Question: What about price?

Answer: Price is important but not paramount. All equipment of quality must bear a price in proportion to the skill, time, experience and work attending their creation and manufacture. Things called expensive when justly estimated are usually the cheapest. Equipment built for cheapness and not for excellence of workmanship and end results is the most frequent and certain cause for dissatisfaction on the part of the customer.

Question: When comparing prices, what should be considered?

Answer: 1 - Does the price quoted include installation by a competent service representative?

2 - Does the price quoted include complete technical instructions by a trained company representative?

3 - Does the manufacturer have an advertised sales price or does the dealer charge any price he may see fit?

4 - Is the trade-in allowance fair or is the dealer boosting the list price in order to allow a greater trade-in allowance?

5 - What accessories are included in the sales price? Are such items as foot switch, cone, filter, timer, stabilizer etc. charged as extras in order to bring the base price down?

6 - What assurance do I have of service after the installation is made?

7 - Is the Company reliable?

8 - How long has the Company been in business?

9 - Will I be able to get service 10 years from now?

The above factors should be given the most serious consideration when contemplating the purchase of equipment. H. G. Fischer & Co. build equipment to fit every price range and every radiographic and fluoroscopic need and is backed up by 39 years of manufacturing know how.

TECHNIC CHART (Use of)

The x-ray technic chart enclosed in this manual is one made up for a Fischer De Luxe 100 X-ray Unit. However, this chart can be adapted to any machine up to and including 100 milliamperes in capacity. For instance, if the machine you desire to use has a capacity of 25 milliamperes employ the time factors shown in the 25 milliampere column and the voltage technic shown on the left hand side of the chart which will give the desired density for the screens and films you are using. If the machine has a capacity of 50 milliamperes employ the time factors shown in the 50 milliampere column. In other words use the time factors shown in the milliampere column corresponding to the milliampere capacity of your machine. The kilovoltage will be the same regardless of what milliampere factor is used. On x-ray units of 100 milliamperes or less high speed screens should be used. Ordinarily the voltage technic shown in voltage technic #2 will give proper density. However, if the screens are old or slow it may be necessary to employ voltage technic #3 or possibly #4 to secure the proper density.

Initial Anode-Film Distance	To Change Focal Distance Factors							
20"	1.0	1.6	2.3	3.2	4.0	5.8	9.0	13.0
25"	.64	1.0	1.4	2.1	2.6	3.7	5.8	8.3
30"	.44	.69	1.0	1.4	1.8	2.6	4.0	5.8
36"	.31	.48	.69	1.0	1.2	1.8	2.8	4.0
40"	.25	.39	.56	.81	1.0	1.4	2.3	3.2
48"	.17	.27	.39	.59	.89	1.0	1.6	2.3
60"	.11	.17	.25	.36	.44	.64	1.0	1.4
72"	.08	.12	.17	.25	.31	.44	.69	1.0
New Anode-Film Distance	20"	25"	30"	36"	40"	48"	60"	72"




Radiographic Technic Chart

DE LUXE MODEL

RF-100 Shockproof X-Ray Apparatus

Exposure Time (sec.)	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
9	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
10	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
11	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
14	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
16	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
17	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
18	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
19	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
20	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1

Exposure Time (sec.)	1		2		3		4		5		6		7		8		9		10		11		12		13		14		15		16		17		18		19		20	
	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2	1	2		
1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
3	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
4	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
5	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
6	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
7	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
8	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
9	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
10	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
11	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
14	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
16	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
17	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
18	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
19	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	
20	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	



Following the horizontal exposure time will be found by following the horizontal exposure time to its intersection with the diagonal line, then read down dropping vertically from the same point to the exposure time.

The exposure time of exposure will be found by following the horizontal exposure time to its intersection with the diagonal line, then read down dropping vertically from the same point to the exposure time.

ALL WITH 100-200 CHARGE BARS OR SETTING WITH FILM AT ROOM TEMPERATURE

H. G. FISCHER & CO.
9431 91 W. BIRMINGHAM AVE. - BIRMINGHAM, ALA.

Fischer

"PRECISION BUILT" X-RAY UNITS Simplify Radiographic Technique

*The unit illustrated is only one of
this extensive line... see others inside*

New
DE LUXE MODEL RF
250 Milliampere
**SHOCKPROOF X-RAY
APPARATUS**




A. G. FISCHER & CO.

Manufacturers

POWERFUL · COMPACT · FLEXIBLE RADIOGRAPHIC-FLUOROSCOPIC X-RAY APPARATUS



Fischer DE LUXE MODEL RF 250 Milliamper X-RAY APPARATUS - For Vertical Fluoroscopy
Automatic control of all factors of technic.



Fischer MODEL L 200 Milliamper X-RAY APPARATUS with Tilt Table
A powerful unit recommended for 45° AP and lateral radiography in the vertical and horizontal positions.

Fischer DE LUXE MODEL RF 100 Milliamper X-RAY APPARATUS with Fluoroscope-Mounted Tilt Table
Only small installation space required.



OVER 100,000 SATISFIED USERS


Throughout the world greatly increase their services to patients with Fischer units. May we send you illustrated circulars showing units that may well meet your requirements? Return of the enclosed card will bring complete details on any Fischer model, and information as to low prices and our liberal payment plan.

High Quality Units of Demonstrated Efficiency

Offering the Value Per Dollar Expended Obtainable in the Industry

Mail Enclosed for Illustrated Literature on Fischer Units Prices, and Our Small Down Payment, Low, Monthly Payments, Income-While-You-Pay Plan

H. G. Fischer & Co.



MODEL TF 30 Milliamper or 50 Milliamper FLUOROSCOPIC X-RAY APPARATUS with Radiographic Facilities


Dependable power and exceptionally fine control. Great efficiency and durability.



Fischer DE LUXE, RAIL MOUNTED, REMOTE CONTROL MODEL TC 30 Milliamper and 50 Milliamper, with Roentgenographic Table
Beautiful, modernly designed X-ray unit for horizontal and vertical radiography - exceptionally fine milliamperage and voltage control.



Fischer MODEL 100 Milliamper X-RAY UNIT With Urological Table
Meets all requirements of modern urological technics. All desired settings of table and X-ray unit easily arranged.



Fischer MODEL TC 30 Milliamper or 50 Milliamper MOBILE X-RAY APPARATUS
Extremely efficient, low priced unit with great flexibility.

Fischer DE LUXE
MODEL 400
SHORT WAVE
DIATHERMY
UNIT
F.C.C. Type
Approval, D-479

A.M.A. acceptance.
Highest quality unit.
Unexcelled performance in all types of
short wave diathermy
applications.



Fischer
MODEL CK
ULTRA
VIOLET
GENERATOR

Powerful Quartz
and Corax
tubes.



Fischer
MODEL J
GALVANIC
AND
CONTRACTILE
CURRENTS
GENERATOR

16 distinct low
voltage currents.
Valuable in di-
agnosis and ther-
apy. Patented
features assure
exact wave form.



ELECTROMEDICAL EQUIPMENT

May we send you information on
these modern units and our liberal
payment plan?

Fischer MODEL 10 SHORT WAVE
DIATHERMY UNIT
F.C.C. Type Approval, D-488

Powerful, low cost, portable. Highly
efficient for condenser pad applications.



Fischer TABLE
MODEL 200
SHORT WAVE
DIATHERMY
UNIT
F.C.C. Type
Approval, D-484

Mobile, portable.
Patented, adjust-
able induction
electrode gives
deep, uniform heat
to all parts of
body.




949 Form 5125A
Printed in U.S.A.

H. G. FISCHER & CO. · FRANKLIN PARK, ILLINOIS

MANUFACTURERS OF X-RAY AND ELECTROMEDICAL EQUIPMENT

9451-9491 W. Belmont Avenue
TELEPHONES: (Franklin Park) Gladstone 3-0612 (Chicago) Tivoli



Deluxe Model RF-250 X-ray Apparatus with Fluoro-
scope Mounted Motor Driven Tilt Table, having
Seat and Shoulder Rest, 17" x 17" 1/2 second 50
line 30" focal distance Bucky, Compression Device,
1/20 second Automatic Reset Timer, Rails, Rollers
and Blocks for Tilt Table, Foot Switch, Cone B
and Filter, with one Stationary Anode Tube under
table and one Rotating Anode Tube over table.

The Deluxe RF-250 X-Ray Apparatus may be had
with Hand Tilted Table or Motor Driven Table.
It may be had with:

1. One stationary anode tube only
2. Two stationary anode tubes
3. One rotating anode tube only
4. One stationary anode tube and one
rotating anode tube



New

DE LUXE MODEL RF

100 Milliampere

SHOCKPROOF X-RAY APPARATUS

H.G. FISCHER & CO.

AUTOMATIC CONTROL • RADIOGRAPHIC-FLUOROSCOPIC COMBINATION

COMBINATION FLUOROSCOPIC RADIOGRAPHIC X-RAY

HORIZONTAL FLUOROSCOPY



DISTINCTIVE IN THE FIELD OF RADIOGRAPHY

The new H. G. FISCHER & CO. De Luxe Model RF 100 Vertigo-type Shockproof Radiographic-Fluoroscopic X-ray Apparatus combines accurate, simple, automatic control of all desired radiographic settings of 5, 10, 25, 75 and 100 Ma. from 30 to 100 ppy. A fully constructed, built-in table permits X-ray and fluoroscopic examinations in all positions from Trendelenburg to vertical. The De Luxe Model RF 100 is acco-


modate features of construction located by more than 30 years leadership in design, development and manufacturing of mobile equipment and accessories in electronic high quality X-ray and fluoroscopic equipment for the last 30 years. The compact, serviceable, constructed De Luxe Model RF 100 is the most efficient, reliable, and most versatile X-ray and fluoroscopic apparatus that can be installed in a small space.

VERTICAL RADIOGRAPHY





Vertical Fluoroscopic



Angular Fluoroscopic

Automatic control of all factors of technic permits radiographic and fluoroscopic examinations with the newly designed FISCHER De Luxe Model RF-100. Milliamperage Shockproof X-ray apparatus to be accomplished easily, quickly and accurately. Moderately priced, flexible and efficient, the FISCHER De Luxe "RF-100" is a worth while investment and one that can be relied upon to pay dividends of excellent results and operating economy, with long, satisfactory, trouble-free service.

THE DELUXE "RF-100" TILTING BUCKY RADIOGRAPHIC-FLUOROSCOPIC TABLE

Delicate balance of the table permits finger-tip operation in all positions. Pressure on one of eight positions, from Trendelenburg to vertical, permits use of the table in 15°. Rail mounting of the table, permitting a lengthwise travel of 5 ft., enhances its flexibility when installed in small quarters.


The fluoroscopic screen staging is mounted on a specially designed carriage for smooth operation on both cross and horizontal travel. Large ball-bearing surfaces provide easy movement and long service.

A special self-aligning guide plate is attached to the fluoroscopic shutter mechanism which permits coupling of the tube to the shutter for fluoroscopic work. Locking of the tube to the shutter mechanism is automatic and self-centering. No locking screws or clamps are necessary. This coupling operation permits changing to the fluoroscopic set-up even in a dark room, and can be accomplished with the table in any position without disturbing the patient.

The 17" x 17" Bucky diaphragm is extremely sturdy in construction and incorporates a 50 line grid with a filter ratio of 6 to 1. An interlocking device automatically centers the tubestand on the Bucky diaphragm longitudinally, permitting the tubestand and Bucky to move in unison the entire length of the table.

A heavy duty tubestand with a manually operated lateral stereoscopic shift of 12 inches is mounted on two parallel rails close to the floor. Provisions are incorporated for leveling the floor rails both laterally and longitudinally by means of leveling screws. The tubestand is so constructed as to permit manipulation from over-the-table to under-the-table positions with a minimum of effort. The sturdy construction of the tubestand and floor rails makes possible the use of either the regular double focus oil immersed tube or the double focus oil immersed rotating anode tube, which can be furnished as optional equipment at additional cost.

A specially constructed heavy duty shockproof oil immersed X-ray tube is provided, together with shockproof cables of sufficient length to connect the tube with the transformer which is normally mounted beside the tubestand. Remote mounting of the transformer is possible by the use of longer cables.



The De Luxe "RF-100" occupies an overall space 6 ft. x 5 ft. 7 in., including standard length rails, table and control mechanism, permitting maximum crosswise travel of tubestand. Minimum height required for tubestand 7 ft. 4 in.

De Luxe Automatic Control

The FISCHER De Luxe "RF-100" control mechanism incorporates many exclusive features which make radiography a simple, precise science. There is no guesswork when using the De Luxe "RF-100" radiographic and fluoroscopic unit. It eliminates as far as humanly possible every electrical and mechanical variable in radiographic technique. Its many automatic compensating devices, constantly in operation, work to the end-result of saving time in handling patients, prolonging tube life and protecting the apparatus from failures due to faulty technique, thus assuring consistently high grade radiographs with resulting film economy.

SPECIAL FEATURES

1. Automatic preselection of milliamperage.
2. Automatic changing of local spots.
3. Automatic thermal induction regulator.
4. Compensated kilovolt meter (illuminated).
5. Double scale milliamperage meter (illuminated).
6. Automatic space charge compensator.
7. Line compensator.
8. Built-in spot change-over device.
9. When doing spot film radiography, both fluoroscopic and radiographic exposures are made with the foot switch. Focal spots are automatically changed to the proper focal spot for the work being done.
10. Fluoroscopic room light control.
11. Overload circuit breaker and main line switch.
12. Separate vernier control of milliamperage for fluoroscopy and therapy.
13. 100 steps of auto-transformer control in 2 KVP steps.
14. Accurate automatic reset timer.
15. Vertical panel gives eye-level view of meters.
16. Five plug receptacle connections are provided for plugging in the various essential accessories. All receptacles are labeled and are of different types, preventing the possibility of plugging into the wrong receptacle. Plug-in connections arranged to accommodate foot switch, spot tunnel, fluoroscopic room light control, Bucky trip cable and therapy timer.

CONTROL FEATURES

Standard Mobile Control on Film Cabinet

Permits convenient location of controls; increases convenience in operation. Provides storage cabinet for films and cassettes. Lead-lined storage bin with slip-proof lock, mounted on mobile base with ball-bearing casters measures 6" x 16" x 20" (inside).

SPECIAL FEATURES

1. Automatic reset timer, calibrated in 1/20 second, maximum 14 seconds.
2. Milliamperage meter has double scale: kilovoltage meter, four scales.
3. Focal selector switch to determine local spot. Pilot lights show when proper focal spot is in use.
4. Main (line) switch, which functions also as a circuit breaker.
5. Main and auxiliary controls for milliamperage.
6. Main and auxiliary controls for kilovoltage; provides 60 points of adjustment.
7. Pre-adjustment switch. Milliamperage pre-settings may be made without energizing the X-ray tube.
8. Timer push button switch.
9. Convenient handles for moving control mechanisms. Receptacles for foot switch and Bucky diaphragm connections are at sides of control mechanism. The position and angle of control panel make for maximum convenience in adjustment and readings. Markings are clear and will remain so permanently.

CONSTRUCTION

Features

OPERATION

DE LUXE CONTROL FEATURES

EQUIPMENT PROTECTION. The main line switch is of the automatic type adjusted to protect the apparatus from short circuit conditions.

LINE COMPENSATOR. The line compensator permits ready adjustment for changes in line voltage from day to day or hour to hour. The line compensator switch provides for a 20 volt variation in service voltage. Adjusting the line compensator switch so that the line voltage is constant assures the same definite radiographic results at all times. This adjustment guarantees precise operation and consistency of results under variable service conditions.

AUTOMATIC PRESELECTION OF MILLIAMPERAGE. The automatic milliamperage selector switch positively and accurately selects any electrical and mechanical factors which are normally constant. It provides five radiographic milliamperage selections: 10, 25, 50, 75 and 100 milliamperes. One very important function of the milliamperage selector switch is the automatic selection of the correct local spot. This prevents accidental overloading of the fine focus X-ray tube. The FISCHER control never forgets to change focal spots. This switching of local spots is indicated by the red and green pilot lights on the control panel.

SPACE CHARGE COMPENSATOR. The function of the space charge compensator is to maintain a constant milliamperage regardless of kilovolt settings. Normally, the milliamperes will increase with change of kilovolt even though the filament current is at a constant value. The action of the space charge compensator is entirely automatic and requires no attention on the part of the operator.

THERMAL INDUCTION REGULATOR. This is an exclusive FISCHER development. The thermal induction regulator compensates the filament circuit for changes in X-ray tube characteristics due to heating. Used in conjunction with a filament station (standard equipment), accurate milliamperage values are assured regardless of whether the tube is cold or hot. Milliamperage values are maintained within 5%.

PRECISE KILOVOLTAGE CONTROL. The De Luxe "RF-100" auto-transformer is a heavy duty type, with more than sufficient capacity to carry the heaviest rated load. Accurate preselection of kilovoltage is obtained throughout the entire voltage range at any milliamperage value. The major and minor auto-transformer switches provide 100 selections of kilovoltage at approximately 2 KVP per step. The major switch provides 10 kilovolts each—the minor switch provides 10 steps of 2 kilovolts each.

FLUOROSCOPIC CONTROL. Separate kilovoltage and milliamperage controls are provided for fluoroscopy and therapy. This is particularly advantageous when doing spot radiography. The milliamperage filament control is of the vernier type and permits accurate control of milliamperage over a range of 2 to 10 Ma., giving precise control for fluoroscopy or therapy.

TIMER. A motor driven timer with a timing range of 1/10 to 14 seconds is provided. Extremely accurate, it incorporates an automatic reset feature permitting stereoscopic films without resetting the timer. Switching arrangements are provided for automatically tripping the Bucky diaphragm and starting the radiographic exposure.

ACCURATE KILOVOLT METER. A single scale kilovolt meter is provided, instead of the usual multiple scale type, which increases the possibility of reading the wrong scale. The meter is compensated to allow for the inherent drop in the high tension transformer which increases as the milliamperage load increases. The kilovolt meter reads actual kilovolts at all times, regardless of the milliamperage selection, and the same radiographic densities will be obtained at all milliamperage values, provided the milliamperage second time is constant.

SPOT FILM CHANGE OVER DEVICE. Ordinarily the spot film change over device is considered extra equipment and is provided at extra cost. In the De Luxe "RF-100" the spot change over device is a built-in feature and is provided as standard equipment. The spot tunnel is extra.

FLUOROSCOPIC ROOM LIGHT CONTROL. A room light control is another built-in feature of the De Luxe "RF-100." No special relays or costly foot switches are required; merely plug into the outlet provided in the control stand.

MOBILE CONTROL FEATURES

Mobile Control on Film Cabinet available on separate arrangement. Control mechanism (see preceding page).

HIGH TENSION TRANSFORMER

Advanced design and sound engineering principles have been incorporated in the De Luxe "RF-100" high tension transformer. The De Luxe "RF-100" transformer, compact in size, is of the closed core type, efficient cooling. Entire transformer assembly is mounted in cast-iron oil and is contained in welded steel tank of heavy gauge steel.

TRANSFORMER CAPACITY

Radiography	Fluoroscopy	Therapy
100 Ma. 96 KVP	5 Ma. 90 KVP	5 Ma. 96 KVP

All capacities shown above represent 85% of the maximum output of the apparatus, 15% being reserved for a safety overload factor.

TUBE CARRIAGE ARM.

A single tube carriage arm supports the X-ray tube and its movements through a supporting sleeve coating on ample ball bearing surfaces, which provides free and easy movement both horizontally and vertically. A tension type mounting lock is used for securing the X-ray tube unit to the tube carriage arm. This provides a flexible type mounting, allowing the rotation of the X-ray tube 360° around the axis of the arm and 20° each way from the center of the lock mounting, so that any required angle may be secured with ease. Angle locks within easy reach are provided for securing tubehead in all positions, horizontally and vertically.

SHOCKPROOF X-RAY TUBE. A specially constructed heavy duty shockproof high tension X-ray tube is provided, together with shockproof cables of sufficient length to connect the tube with the transformer, which is mounted inside the tube stand.


SHOCKPROOF THROUGHOUT. The De Luxe "RF-100" is rendered completely shockproof by means of an all insulated, shockproof double focus tube installed in a grounded metal tube housing. The tube housing is connected with the high tension transformer through high tension cables having a grounded metal sheath. Positive grounding of the entire unit is accomplished through a ground wire incorporated in the line cable.

FLUOROSCOPIC SCREEN AND MOUNTING. A Permaplan Type "B-3" fluoroscopic screen, 12" x 18" is provided on regular equipment. The screen automatically centers on the tube in all positions and is supplied with a spring counterbalance for use in the horizontal position. Spring tension can be released when the tube is used in the vertical position. Locks are provided for locking the screen and tube in any position crosswise or lengthwise of the table. A lock is also furnished for locking the screen rigidly at any desired height from the table top.

BUCKY DIAPHRAGM. The latest model (36 line grid with a ratio of 6 to 1) Bucky diaphragm, fully counterbalanced in all positions of the table, with a timing range of 1/10 to 30 seconds Bucky diaphragm as standard equipment. A 1/10 to 30 seconds Bucky diaphragm can be supplied if desired. An interlocking device holds the diaphragm and tubehead together so that they move at the same time. Bucky and tubehead disengage tubestand when the Bucky is not in use. Simple movement disengages tubestand when the Bucky is not in use.

FINISH. The De Luxe Model "RF-100" is exceptionally attractive in hammered tone gray with polished chrome trim.

Cat. No.	Code
5100	FISCHER De Luxe Model RF-100 Milliamperage Shockproof X-ray Apparatus complete with De Luxe "RF-100" Automatic Control and Timer, Fluoroscopic Mounting, Tilt Table and Bucky unit, including essential accessories listed below.
5099	FISCHER De Luxe Model RF-100 Milliamperage Shockproof X-ray Apparatus complete with Standard Mobile Control on Film Cabinet, Fluoroscopic Mounting, Tilt Table and Bucky unit, including essential accessories listed below.
	*Essential Accessories: (1) 100 Ma. Foot Switch, (2) 100 Ma. Foot Switch, (3) 100 Ma. Foot Switch, (4) 100 Ma. Foot Switch, (5) 100 Ma. Foot Switch, (6) 100 Ma. Foot Switch, (7) 100 Ma. Foot Switch, (8) 100 Ma. Foot Switch, (9) 100 Ma. Foot Switch, (10) 100 Ma. Foot Switch.
	Operating Voltage: 110-240 Volts, 50-60 Cycles.
	Max. Current: 150 Amperes.
	Power Rating: 15 KVA.
	Shipping Weight: Domestic 2650 lbs., Foreign 3300 lbs.



NEW HOME OF H. G. FISCHER & CO.

Representative Units of the Complete Line of H. G. FISCHER & CO.

The complete H. G. FISCHER & CO. line includes many models of Shockproof X-ray apparatus (medical and dental), Short Wave Diathermy Units, Galvanic and Contractile Currents Generators, Ultra Violet Generators, Infra Red Generators, other apparatus, supplies and accessories. H. G. FISCHER & CO. has always maintained the policy of meeting or exceeding the standards of any competing units designed for the same purpose, regardless of price, and pricing all apparatus at the lowest point consistent with quality manufacture.

SINCE 1910, H. G. FISCHER & CO. has been one of the largest manufacturers of Shockproof X-ray and Electromedical equipment. Pioneers in the field of X-ray and Electromedical equipment, H. G. FISCHER & CO. is recognized throughout the world today for apparatus of dependable service and great durability. H. G. FISCHER & CO. is the ORIGINAL COMPANY. It has no connection with any other company having a similar name. No H. G. FISCHER & CO. equipment is manufactured outside of the Chicago area.

More than thirty-eight years of experience in engineering, design and construction of X-ray and Electromedical apparatus, together with fine workmanship and materials, are built into all H. G. FISCHER & CO. apparatus. The fact that today over 100,000 physicians, dentists, hospitals, clinics, universities and other medical organizations in the United States and throughout the world use H. G. FISCHER & CO. apparatus reflects its outstanding quality and service.

All H. G. FISCHER & CO. apparatus is serviced and sold through factory-trained representatives who are well qualified to render necessary technical assistance. Our comparatively small overhead expense and factory-trained representation enable us to give purchasers such value per dollar expended as cannot be excelled by any other manufacturer in the field.

The new home of H. G. FISCHER & CO. is now conveniently located at 9451-91 W. Belmont Ave., Franklin Park, Illinois, a western suburb of Chicago (3209 North and 9400 West, by the Chicago numbering system). Doctors are cordially invited to visit the new H. G. FISCHER & CO. plant, where modern X-ray and Electromedical equipment will be displayed and demonstrated. A visit entails no obligation.

H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS

1943 Form 5115

Printed in U.S.A.



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus




Model RF-100 100 Milliamperes Shockproof X-ray Apparatus



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus



Model RF-100 100 Milliamperes Shockproof X-ray Apparatus

The DeLuxe RF-100 X-Ray Apparatus may be had with Hand Tilted Table or Motor Driven Table. It may be had with:

1. One stationary anode tube only
2. Two stationary anode tubes
3. One rotating anode tube only
4. One stationary anode tube and one rotating anode tube



"SPACESAVER"

Radiographic-Fluoroscopic Unit
and Examining Table
100 or 250 Milliamperes




HORIZONTAL RADIOGRAPHY

Any part of a patient's body can easily be radiographed in the horizontal position. The examining table is 26 1/4 inches wide and 73 inches long including foot rest—capable of accommodating any patient. The tubehead travels up and down and across the cabinet, giving a complete range of operation. Either a standard Bucky diaphragm or a stationary grid may be used.


H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS
(Suburb of Chicago)




VERTICAL FLUOROSCOPY

The table top is raised and locked in the vertical position, forming a thoroughly adequate vertical fluoroscopic panel for the patient to lean against. The tubehead is then rotated on its horizontal axis 50° from the radiographic position. The 12 by 16 inch fluorescent screen is automatically centered with the tubehead. The fluoroscopic shutters are attached to the tubehead so that at all times, in either the horizontal or vertical position, maximum protection is assured.



The rigid, sturdy, channel construction details are shown in this rear view of the FISCHER "Spacesaver".



Floor plan view with dimensions of the FISCHER "Spacesaver", 100 Ma or 250 Ma. Scale 1/2 inch = 1 foot.

Cat. No.	DESCRIPTION	Code
5188	FISCHER "Spacesaver 250" with De Luxe Automatic Control, 1/20 second Automatic Reset Timer and Foot Switch HOCOH	
5186	FISCHER "Spacesaver 100" with De Luxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch HELMS	
5187	FISCHER "Spacesaver 100" with Standard Mobile Control, 1/10 second Motor-driven Synchronous Timer and Foot Switch	HEXAR
Following are suggested accessories not included in above prices.		
5172	20-line grid 14" x 17" 1/2 second 30" focal distance Bucky with magnetic release, x-ray timing contacts, rollers and Bucky connecting cord	HAGUS
4736	Cone	TIPPS
5175	Stirrups	HAINE
5176	Pad for Table Top	HAITT
5177	30" focal distance stationary aligned grid—size 14"x17"	HADAN
5178	Moveable support and tray for cassette and stationary grid	HADAT




The table top is compound hinged at one end and can easily be raised to the vertical position, where it automatically locks.

VERTICAL RADIOGRAPHY

The tubehead is turned and focused on a patient standing against a wall-mounted cassette holder at the correct distance. Vertical radiographs can be made at either side of the table.





VERTICAL FLUOROSCOPY

The table top is raised and locked in the vertical position, forming a thoroughly adequate vertical fluoroscopic panel for the patient to lean against. The tube head is then rotated on its horizontal axis 90° from the radiographic position. The 12 1/2 inch fluoroscopic screen is automatically centered with the tube head. The fluoroscopic shutters are attached to the tubehead so that at all times, in either the horizontal or vertical position, maximum protection is assured.

FISCHER "Spacesaver"

Radiographic-Fluoroscopic Unit and Examining Table, 100 Ma or 250 Ma

Both horizontal and vertical radiography and fluoroscopy.
Minimum floor space.
Minimum operating effort.
Absolute safety for operator and patient.

The FISCHER "Spacesaver" Combination Radiographic-Fluoroscopic Unit and Examining Table is the latest development in the FISCHER complete line of modern, shockproof x-ray equipment. This new line is distinguished for outstanding performance and exceptional durability. As its name implies, the "Spacesaver" is the most compact of all x-ray equipment of such power and versatility. It was designed to occupy a minimum amount of floor space and to give dependable performance with the least effort on the part of the operator. It may be used at any dry examining table.

The great breadth of its uses is as convincing as the ease with which it functions. It can be switched from position for horizontal fluoroscopic work to horizontal radiographic work and vice versa by a simple swing of the tubehead. The change to vertical fluoroscopic and vertical radiographic positions is equally as easily accomplished.

A Panatone Type R-2 full size 12" x 16" fluoroscopic screen is included with the FISCHER "Spacesaver" at no additional cost. The size and quality of the screen is a feature of the FISCHER "Spacesaver" fluoroscopic screen safety.

The FISCHER "Spacesaver" fluoroscopic screen safety is a feature of the FISCHER "Spacesaver" fluoroscopic screen safety.

"Spacesaver 100" capacity is 100 Ma at 90 KVP, 75 Ma at 85 KVP, 50 Ma at 80 KVP, 25 Ma at 75 KVP, 10 Ma at 70 KVP, and 5 Ma at 65 KVP.

"Spacesaver 250" capacity is 250 Ma at 90 KVP, 200 Ma at 100 KVP, 100 Ma at 120 KVP, 10 Ma at 120 KVP, intermittent, and 7 Ma at 120 KVP, continuous.

Finish is attractive, lasting, hammered gray.

DE LUXE AUTOMATIC CONTROL

The FISCHER De Luxe Automatic Control mechanism incorporates many exclusive features which make radiography a simple, precise science. There is no guesswork when using the FISCHER "Spacesaver". It eliminates all possible every electrical and mechanical variable in radiographic technique. Its many automatic compensating devices, constantly in operation, work to the end result of saving time in handling patients, prolonging tube life and protecting the apparatus

from failures due to faulty technique, thus resulting in consistently high quality radiographs with resulting film economy.

SPECIAL FEATURES OF DE LUXE AUTOMATIC CONTROL

1. Automatic pre-rotation of milliamperes.
2. Automatic changing of load film.
3. Compensated milliamperes meter (standard).
4. Double scale milliamperes meter (optional).
5. Automatic space charge compensation.
6. Line compensator.
7. Built-in space charge over control.
8. When doing spot film radiography, both fluoroscopic and radiographic exposures are made with the foot switch. First spot are automatically changed to the proper level spot for the work being done.
9. Fluoroscopic room light control.
10. Overload circuit breaker and auto line switch.
11. Separate vertical control of milliamperage for fluoroscopy and therapy.
12. 150 steps of auto-transformer control in 2 KVP steps.
13. Automatic automatic timer from 1/125 sec. to 1/2 sec. for 100 Ma "Spacesaver" and 1/10 sec. for 100 Ma "Spacesaver".
14. Vertical panel gives eye level view of patient.
15. Five play magnetic condensers are provided for protection in preventing the possibility of playing into the wrong way table. Plug-in connections arranged to accommodate foot switch, spot timer, fluoroscopic room light control, back up cable and therapy time.

STANDARD CONTROL

The FISCHER Standard Control permits convenient location of controls and increases convenience in operation.

SPECIAL FEATURES

1. Motor-driven synchronous timer, calibrated in 1/125 second, maximum 14 seconds.
2. Milliamperes meter has double scale kilovoltage meter, four scales.

3. Foot selector switch to determine load spot. Foot light shows when proper load spot is in use.
4. Main (line) switch, which functions also as a circuit breaker.
5. Main and auxiliary controls for milliamperes.
6. Main and auxiliary controls for kilovoltage provides 90 points of adjustment.
7. Pre-adjustment switch. With ampere positionings may be made without changing the X-ray tube.
8. Timer push button switch.
9. Rectifier for foot switch and back up discharge connections. The position and mode of control panel make for maximum convenience in adjustment and readings. Markings are clear and will remain so permanently.




HORIZONTAL FLUOROSCOPY


The changing from the horizontal radiographic to the horizontal fluoroscopic position, or the reverse, is made without moving patient from table. This feature is a specially advantageous when, for example, after setting a broken bone with the aid of the fluoroscopic, it is desired to radiograph the results in seconds. The fluoroscopic field is 44 inches long and may be utilized in 10 seconds.

FISCHER & CO.





The rigid, sturdy, channel construction details are shown in this rear view of the FISCHER "Spacesaver".



Floor plan view with dimensions of the FISCHER "Spacesaver", 100 Ma or 250 Ma. Scale 1/2 inch = 1 foot.



The table top is compound hinged at one end and can easily be raised to the vertical position, where it automatically locks.

VERTICAL RADIOGRAPHY

The tubehead is turned and focused on a patient standing against a wall-mounted cassette holder at the correct distance. Vertical radiographs can be made at either end or either side of the table.



Cat. No.	DESCRIPTION	Code
5185	FISCHER "Spacesaver 250" with De Luxe Automatic Control, 1/20 second Automatic Reset Timer and Foot Switch. HOCOH	
5186	FISCHER "Spacesaver 100" with De Luxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch. HELMS	
5187	FISCHER "Spacesaver 100" with Standard Mobile Control, 1/10 second Motor-driven Synchronous Timer and Foot Switch. HEXAR	
Following are suggested accessories not included in above prices.		
5172	20-line grid 14" x 17" 1/2 second 30" focal distance Bucky with magnetic release, x-ray timing controls, rollers and Bucky connecting cord. HAGUS	
4796	Cone. TIPS	
5175	Straps. HAINE	
5176	Pad for Table Top. HAITT	
5177	30" focal distance stationary aligned grid - size 14"x17". HADAN	
5178	Movable support and tray for cassette and stationary grid. HADAT	



NEW HOME OF H. G. FISCHER & CO.

United States Navy



BUREAU OF MEDICINE AND SURGERY
Certificate of Achievement

H. G. Fischer & Co.

In grateful acknowledgment of exceptional cooperation and outstanding services rendered to the Medical Department of the Navy during the period of World War II

John H. Fisher
Chief, Bureau of Medicine and Surgery



H. G. FISCHER & CO.
FRANKLIN PARK, ILLINOIS

Manufacturers of high quality X-ray and electromedical equipment since 1915.

Holders of the U. S. Army-Navy "E" Award with three additional awards represented by the stars on the flag, and holders of the U. S. Navy Certificate of Achievement - all for outstanding services rendered.

This series of awards constitutes the most visible recognition granted any manufacturer of X-ray equipment.

Producing - in a new \$2,000 sq. ft. factory - efficient, durable, attractive apparatus, which offers customers the greatest value per dollar expended obtainable in the entire industry.



H. G. Fischer & Co. receives U. S. Navy Certificate of Achievement from Chief, Bureau of Medicine and Surgery, John H. Fisher.


OVER 100,000 SATISFIED USERS THROUGHOUT THE WORLD

H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS

1249 Form 3130

PRINTED IN U. S. A.



NEW HOME OF H. G. FISCHER & CO.

United States Navy

BUREAU OF MEDICINE AND SURGERY
Certificate of Achievement

H. G. Fischer & Co.

In grateful acknowledgment of exceptional cooperation and outstanding services rendered to the United Department of the Navy during the period of World War II



H. G. FISCHER & CO.
FRANKLIN PARK, ILLINOIS

Manufacturers of high quality X-ray and electro-medical equipment since 1910.

Holders of the U. S. Army-Navy "E" Award with three additional awards represented by the stars on the flag, and holder of the U. S. Navy Certificate of Achievement — all for outstanding services rendered.

This series of awards constitutes the most signal recognition granted any manufacturer of X-ray equipment.

Producing — in a new \$2,000,000 factory — efficient, durable, attractive apparatus, which offers customers the greatest value per dollar expended obtainable in the entire industry.



A. M. Deering, Rear Admiral, MC, U. S. Navy, presenting Certificate of Achievement to A. M. Deering, President, H. G. Fischer & Co. Official U. S. Navy photograph.

OVER 100,000 SATISFIED USERS THROUGHOUT THE WORLD


H. G. FISCHER & CO.
9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS

1249 Form 3190

PRINTED IN U. S. A.



"SPACESAVER 30"



Radiographic-Fluoroscopic Unit and Examining Table

H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS

- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus
- Model "T" 30 or 50 Million Per Second X-ray Apparatus

FISCHER "Spacesaver 30"

Low in Price
High in Quality
Greatest Value per Dollar Expended
Proven Dependability

RADIOGRAPHIC-FLUOROSCOPIC UNIT AND EXAMINING TABLE

The FISCHER "Spacesaver 30" Combination Radiographic-Fluoroscopic Unit and Examining Table is an outstanding development in the FISCHER complete line of modern, shockproof x-ray equipment. This great line is distinguished for exceptional performance and great durability. The "Spacesaver 30" is the most compact of all x-ray apparatus of such power and versatility. It was designed to occupy a minimum amount of floor space and to give dependable performance with the least effort on the part of the operator. It may be used as an every day examining table as well as an x-ray apparatus. In approximately the same space required for an ordinary examining table you can, in the case of the "Spacesaver 30," have not only an examining table but also a complete x-ray plant.

IDEAL FOR GENERAL PRACTITIONER

The FISCHER "Spacesaver 30" is ideal for the general practitioner. The great range of its uses is as amazing as the ease with which it functions. It can be switched from positions for horizontal fluoroscopic work to positions for horizontal radiographic work, and vice-versa, by a simple swing of the tubehead. The change to vertical fluoroscopic and vertical radiographic positions is equally as easily accomplished.

121 KILOVOLTAGE REGULATIONS

The control is compactly mounted on the cabinet. It has 121 steps of kilovoltage regulations, making possible the universally recognized, valuable, thickness-of-part technique for most accurate radiographic results. A control with this large number of steps of kilovoltage regulation is rarely found in other than the most expensive equipment. This is another Extra Value feature offered in the FISCHER "Spacesaver 30."

The control is so constructed that when the Bucky diaphragm is used the timing of the x-ray exposure is done by the x-ray timer and not by the less accurate timing mechanism incorporated in all Bucky diaphragms.

VERTICAL RADIOGRAPHY—The tubehead is turned and aimed at a patient standing against a well-mounted stationary Bucky in the vertical position. Vertical radiographs can be made of either end or side view of the table.

CONSTRUCTION Features OPERATION

MECHANICAL

PATIENT NOT DISTURBED when changing from horizontal fluoroscopy to horizontal radiography. Tubehead is easily swung out from under table, raised and swung to radiographic position over table.

TABLE doubles as every-day examining table and x-ray table. Top is compound hinged and raises easily to function as vertical fluoroscopic panel. Struts are available for use on the examining table. Table dimensions are 26 1/4" wide and 73" long.

TUBEHEAD travels up and down and across the cabinet giving complete range of operation. It has a cross travel shift feature for making stereoscopic views. Housing is leakproof seamless drawn steel. Swinging arm supporting tubehead and fluoroscopic screen is counterbalanced and automatically locks in fluoroscopic and radiographic positions.

VERTICAL TUBE COLUMN for tubehead rotates 360°. Carriage cross members are solid steel bars to eliminate sagging. Eccentric bearings used where necessary to permit adjustment to eliminate play.

FULL-SIZE FLUOROSCOPIC SCREEN, 12" x 16" Patterson Type B-2, has 3/4" lead glass protection. Shutter controls are on the screen stopping. Screen automatically centers with tubehead in fluoroscopic position.

STANDARD BUCKY DIAPHRAGM size 14" x 17", may be used. Bucky is mounted on ball bearings and travels full length of cabinet. Stationary grid may also be used. Bucky or grid tray can be moved out of fluoroscopic field under observation.

CABINET is of rigid, sturdy, channel construction. "Spacesaver 30" control can be mounted on the front or on the back of the cabinet. "Spacesaver 250", "Spacesaver 100", and "Spacesaver 50" all have remote control. Leveling devices permit exact leveling of entire unit.

ELECTRICAL

HIGH OUTPUT on adequate lines: "Spacesaver 30" capacity is 30 MA at 74 KVP, 25 MA at 80 KVP, 20 MA at 85 KVP, 5 MA at 96 KVP—all useful kilovolages.

WALL OUTLET ADEQUATE for operation of "Spacesaver 30" on average office lighting circuit up to 20 MA at 85 KVP. It operates on 110V or 220V, 50-60 cycles.

121-STEP CONTROL of kilovoltage regulation permits the universally used thickness-of-part technique for most accurate radiographic and results. Changes of less than two kilovolts each are possible. This is an Extra Value feature at no additional cost.

EXACT KILOVOLTS register on accurate meter for all milliamperage values.

HEAVY-DUTY CONTACTOR breaks transformer primary current.

TRANSFORMER CORE construction makes it possible to reduce size of tubehead with increased efficiency.

ACCURATE TIMER, 1/4 second to 12 seconds, with built-in contacts automatically trips Bucky and times exposure. The x-ray timer is much more reliable than the less accurate Bucky timing mechanism.

FOOT SWITCH actuates heavy-duty contactor. Neon light on foot switch to facilitate locating it in dark room while making fluoroscopic examinations.

PLUG CONNECTIONS on control for plugging in foot switch, timer and x-ray head inter-connecting cable.

VOLTAGE TAPS on control permit operation of unit on 110V, 120V, 220V, or 240V, 50-60 cycle lines.

X-RAY PROTECTION at tubehead in accordance with standards adopted by International Radium and X-ray Commission and is equivalent to 1.5 mm of lead.

3 EXPANSION CHAMBERS to allow for oil expansion.

PILOT LIGHT indicates when control is energized.

Doctor

May we ask you to CONSIDER

1. How an X-ray would increase your services to your patients.
2. What a great convenience it would be to them and to you.
3. How safe, compact, durable and easily operated today's models are.
4. How an X-ray produces a substantial income.
5. How little cash outlay is required to install X-ray equipment which almost always is more than a self-liquidating investment.

Cat. No.	DESCRIPTION	Code
5188	FISCHER "Spacesaver 250" with Deluxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch.	NOCOR
5186	FISCHER "Spacesaver 100" with Deluxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch.	HELM
5187	FISCHER "Spacesaver 100" with Standard Mobile Control, 1/10 second Automatic Reset Timer and Foot Switch.	HEXAR
5185	FISCHER "Spacesaver 50" including remote control, 1/10 second, synchronous and foot switch.	HAFAN
5184	FISCHER "Spacesaver 30" with Foot Switch.	HADY
Following are suggested accessories:		
5172	20 line grid 14" x 17" 1/16 second 30" focal distance Bucky with negative relative x-ray timing controls, rollers and Bucky connecting cord.	RAGUR
4088	Times and Cable for "Spacesaver 30"	TAKER
4798	Cable	TAPER
5175	Storage	RAINE
5176	Post for Table Top.	HAJIT
5177	30" focal distance stationary aligned grid—size 14" x 17"	HADAN
5178	Movable support and tray for cassette and stationary grid.	HADAT

Prices subject to change without notice. F.O.B. Franklin Park, Ill.

CONSTRUCTION Features OPERATION

MECHANICAL

PATIENT NOT DISTURBED when changing from horizontal fluoroscopy to horizontal radiography. Tubelens is easily swung out from under table, raised and swung to radiographic position over table.

TABLE doubles as every-day examining table and x-ray table. Top is compound hinged and rotates easily to function as vertical fluoroscopic panel. Straps are available for use on the examining table. Table dimensions are 36 1/4" wide and 72" long.

TUBEHEAD travels up and down and across the cabinet giving complete range of operation. It has a cone beam shift feature for making stereoscopic views. Housing is leakproof seamless drawn steel. Swinging arm supporting tubelens and fluoroscopic screen is counterbalanced and automatically locks in fluoroscopic and radiographic positions.

VERTICAL TUBE COLUMN for tubelens rotates 360°. Carriage cone members are solid steel bars to eliminate sagging. Eccentric bearings used where necessary to permit adjustment to eliminate play.

FULL-SIZE FLUOROSCOPIC SCREEN, 12" x 16" Patterson Type B-2, has 1/4" lead glass protection. Shutter controls are on the screen staging. Screen automatically centers with tubelens in fluoroscopic position.

STANDARD BUCKY DIAPHRAGM size 14" x 17", may be used. Bucky is mounted on ball bearings and travels full length of cabinet. Stationary grid may also be used. Bucky or grid tray can be moved out of fluoroscopic field under observation.

CABINET is of rigid, sturdy, channel construction. "Spacover 30" control can be mounted on the front or on the back of the cabinet. "Spacover 250", "Spacover 100", and "Spacover 50" all have remote control. Levelling devices permit exact levelling of entire unit.

Doctor

May we ask you to CONSIDER

1. How an X-ray would increase your services to your patients.
2. What a great convenience it would be to them and to you.
3. How safe, compact, durable and easily operated today's models are.
4. How an X-ray produces a substantial income.
5. How little cash outlay is required to install X-ray equipment which almost always is more than a self-liquidating investment.

ELECTRICAL

HIGH OUTPUT on adequate lines: "Spacover 30" capacity is 30 MA at 74 KVP, 25 MA at 80 KVP, 20 MA at 85 KVP, 5 MA at 36 KVP—all useful kilovolts.

WALL OUTLET adequate for operation of "Spacover 30" on average office lighting circuit up to 30 MA at 85 KVP. It operates on 110V or 220V, 50-60 cycles.

275-STEP CONTROL of kilovoltage regulation permits the universally used thickness-of-part technique for most accurate radiographic end results. Changes of less than two kilovolts each are possible. This is an Extra Value feature at no additional cost.

EXACT KILOVOLTS register on accurate meter for all milliamperes values.

HEAVY-DUTY CONTACTOR breaks transformer primary current.

TRANSFORMER CORE construction makes it possible to reduce size of tubelens with increased efficiency.

ACCURATE TIMER, 1/4 second to 12 seconds, with built-in contacts automatically trips Bucky and times exposure. The x-ray timer is much more reliable than the less accurate Bucky timing mechanism.

FOOT SWITCH actuates heavy-duty contactor. Neon light on foot switch to facilitate locating it in dark room while making fluoroscopic examinations.

PLUG CONNECTIONS on control for plugging in foot switch, timer and x-ray head inter-connecting cable.

VOLTAGE TAPS on control permit operation of unit on 110V, 120V, 220V, or 240V, 50-60 cycle lines.

X-RAY PROTECTION at tubelens in accordance with standards adopted by International Radium and X-ray Commission and is equivalent to 1.5 mm of lead.

3 EXPANSION CHAMBERS to allow for oil expansion.

PILOT LIGHT indicates when control is energized.

Cat. No.	DESCRIPTION	Code
5148	FISCHER "Spacover 250" with Deluxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch.	ROOCH
5146	FISCHER "Spacover 100" with Deluxe Automatic Control, 1/10 second Automatic Reset Timer and Foot Switch.	HELMG
5147	FISCHER "Spacover 50" with Standard Rotary Control, 1/10 second Automatic Reset Timer and Foot Switch.	HEKAR
5145	FISCHER "Spacover 30" including remote control on mobile stand, synchronous and foot switch.	KAFAP
5144	FISCHER "Spacover 30" with Foot Switch.	KADAT
Following are suggested accessories		
5171	20-line grid 14" x 17" 1/2 second 30" fixed distance Bucky with magnetic release, x-ray timing mechanism, rollers and Bucky connecting cord.	BAOCH
4088	Timer and Cable for "Spacover 30".	TAGER
4736	Cone.	TQPS
5175	Straps.	BAINE
5174	Pod for Table Top.	KADTT
5177	30" fixed distance stationary aligned grid—size 14" x 17".	KADAM
5178	Removable support and tray for cassette and stationary grid.	KADAT

Prices subject to change without notice. F.O.B. Franklin Park, Ill.



NEW HOME OF H. G. FISCHER & CO.

United States Navy



BUREAU OF MEDICINE AND SURGERY
Certificate of Achievement

H. G. Fischer

In grateful acknowledgment of exceptional cooperation and outstanding service rendered to the United Department of the Navy during the period of World War II.



H. G. FISCHER & CO.
FRANKLIN PARK, ILLINOIS

Manufacturers of high quality X-ray and electro-medical equipment since 1912.

Holders of the U. S. Army Patent "E" Award with three additional awards represented by the stars on the flag and holder of the U. S. Navy Certificate of Achievement—all for outstanding service rendered.

This series of awards constitutes the most signal recognition granted any manufacturer of x-ray equipment.

Producing—in a new \$2,000 up, its factory—efficient, durable, effective apparatus, which often consumes the greatest value per dollar expended elsewhere in the same industry.

OVER 100,000 SATISFIED USERS THROUGHOUT THE WORLD

H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS

2149 Tenth 5128A

PRINTED IN U. S. A.



RECOMMENDED DARK ROOM AND RADIOGRAPHIC ACCESSORIES
REQUIRED FOR GENERAL RADIOGRAPHY

THIS LIST CAN BE USED WHEN A QUICK SUMMARY IS REQUIRED

DARK ROOM ACCESSORIES

<u>Cat. No.</u>	<u>Amount</u>	<u>Description</u>	<u>Unit</u>	<u>Price</u>	<u>Total</u>
1927	1	Model A 3-Compartment 5 Gal. Buckite Developing Tank			
1717	2	Developing Hangers 8" x 10"	each	\$55.00	\$55.00
1714	2	Developing Hangers 14" x 17"	"	2.10	4.20
3296	1	Floating Thermometer	"	2.25	4.50
4730	1	14" x 17" Cassette equipped with--	"	2.00	2.00
2871	1	Pair 14" x 17" High Speed Screens	"	21.45	21.45
4727	1	8" x 10" Cassette equipped with--	pair	45.20	45.20
2862	1	Pair 8" x 10" High Speed Screens	each	14.85	14.85
1680	2	Cardboard Exposure Holders	pair	15.20	15.20
			each	.55	1.10
3064	1	5 Gal. Bottle Concentrated Developer (Liquid)			
3065	1	1 Gal. Bottle Concentrated Replenisher (Liquid)	each	4.50	4.50
3066	1	5 Gal. Bottle Concentrated Fixer (Liquid)	"	1.25	1.25
			"	4.50	4.50
4765	1	Kodak Adjustable Dark Room Lamp	"	7.58	7.58
4165	1	10-Watt Bulb for above (bulb not incl.)	"	.25	.25
4837	1	Model S Minuteminder Dark Room Timer	"	6.50	6.50
3093	1	Pkg. of 25 - 8 x 10 Films	"	6.10	6.10
3096	1	Pkg. of 25 - 14 x 17 Films	"	17.50	17.50
4847A	1	Film Storage Chest (For Cassettes and Films)	"	32.00	32.00

RADIOGRAPHIC ACCESSORIES

4695	1	Film Marker with Name and Case	each	7.50	7.50
4364	1	Thickness Caliper	"	2.50	2.50
4786	1	Fischer Wall Mounted Cassette holder	"	73.00	73.00
3786	1	14 x 17 Illuminator (Fluorescent)	"	25.00	25.00

NOTE - One Type B (6") Cone is supplied as standard equipment on the Deluxe 250, Deluxe 100 - all T units - MPA-20 and the MPA-30 and will not have to be included in orders for above units. When ordering any of the above units the following cones should be ordered in addition to the Type B cone furnished as standard equipment.

4736	1	Type C Cone (4")	each	8.50	
4735	1	Type D Cone (2½")	"	6.50	

When ordering a Spacesaver 30, Spacesaver 50,
Spacesaver 100 or TF Unit INCLUDE THE FOLLOWING
CONES:

<u>Cat. No.</u>	<u>Amount</u>	<u>Description</u>		<u>Unit Price</u>	<u>Total Price</u>
4737	1	Type B Cone (6")	each	\$ 9.50	
4736	1	Type C Cone (4")	"	8.50	
4735	1	Type D Cone (2½")	"	6.50	

NOTE - On all Spacesaver 50 Units include the
Type J Cone listed below. In addition the Types
B, C, and D Cones are recommended.

5174	1	Type J Cone (10")	each	14.00	
------	---	-------------------	------	-------	--

NOTE - Above cones required Cat. No. 4742 adaptor
if used on TF Units (see optional selection list
attached for price and catalog number)

TOTAL PRICE LESS CONES \$359.78

NOTE - If Fluoroscopy is contemplated the follow-
ing fluoroscopic accessories should be included:

385A	1	Lead Rubber Apron	each	\$15.00	\$15.00
4053	1	Pair Lead Rubber Gloves	pair	15.50	15.50
4842	1	Pair Fluoroscopic Goggles for Eye Accommodation	pair	2.75	2.75

TOTAL \$33.00
FLUOROSCOPIC ACCESSORIES

REQUIRED FOR GENERAL RADIOGRAPHY WITH MINIMUM REQUIREMENTS RECOMMENDED

DARK ROOM

<u>Cat. No.</u>	<u>Description</u>	<u>Unit</u>	<u>Price</u>	<u>Minimum Quantities Recommended</u>
1927	Model A 3-Compartment 5 Gal. Developing Tank (Buckite)		\$55.00)	
	OR)	
2908	Wolf 3-Compartment 5 Gal. Developing Tank		60.00)	
	OR)	1
2909	Model B 3-Compartment 2½ Gal. Developing Tank (Buckite)		44.00)	
	OR)	
2760	Wolf 3-Compartment 2½ Gal. Developing Tank		50.00)	
1717	Developing Hangers 8" x 10"	each	2.10	2
1714	Developing Hangers 14" x 17"	"	2.25	2
3296	Floating Thermometer	"	2.00	1
4730	14" x 17" Cassette equipped with	"	21.45	1
2871	Pair 14" x 17" High Speed Screens	pair	45.20	1
4727	8" x 10" Cassette equipped with	each	14.85	1
2862	Pair 8" x 10" High Speed Screens	pair	15.20	1
1680	Cardboard Exposure Holders	each	.55	2

CHEMICALS REQUIRED FOR 5 GAL. DEVELOPING TANKS

3064	5 Gal. Bottle Concentrated Developer (Liquid)	each	4.50	1
3065	1 Gal. Bottle Concentrated Replenisher (Liquid)	each	1.25	1
3066	5 Gal. Bottle Concentrated Fixer (Liquid)	"	4.50	1
	OR			
685	5 Gal. Pkg. Powdered Developer	each	3.50	1
681	1 Gal. Pkg. Powdered Replenisher	"	1.00	1
708	5 Gal. Can Powdered Fixer	"	3.25	1

CHEMICALS REQUIRED FOR 2½ GAL. DEVELOPING TANK

681	1 Gal. Pkg. Powdered Developer	each	.85	3
707	1 Gal. Pkg. Powdered Fixer	"	.75	3

<u>Cat. No.</u>	<u>Description</u>		<u>Unit Price</u>	<u>Minimum Quantities Recommended</u>
5167	Fluoroscopic Foot Switch for use with T Units	each	\$18.00	1
4930	Fluoroscopic Foot Switch for use with MPA 20	"	30.00	1
5189	Fluoroscopic Foot Switch for use with MPA 30	"	16.50	1
4695	Film Marker with name, lead figures and case (Neon lighted)	"	7.50	1

X-RAY ACCESSORIES

CASSETTE HOLDERS

	LIST PRICE
Fischer Wall-Mounted Tilting Cassette Holder.....	\$ 70.00
Fischer Wall-Mounted Cassette Holder.....	73.00

HAND FLUOROSCOPES

5" x 7" Patterson Shielded Hand Fluoroscope.....	\$ 66.00
5" x 7" Patterson Shielded Operating Fluoroscope.....	79.00
5" x 7" Fischer Hand Fluoroscopes.....	42.00

PATTERSON TYPE "B2" FLUOROSCOPIC SCREENS

5" x 7" Screen.....	\$ 14.00
8" x 10" Screen.....	32.00
10" x 12" Screen.....	48.50
11" x 14" Screen.....	61.50
12" x 16" Screen.....	77.50
14" x 17" Screen.....	95.50

STATIONARY FILTER GRIDS

Parallel Lead Strips Grids

5" x 7"	\$ 35.00
8" x 10"	60.00
10" x 12"	80.00
11" x 14"	95.00
14" x 17"	122.50

36-Inch Focal Distance Aligned Grids

10" x 12"	\$ 95.00
11" x 14"	110.00
14" x 17"	137.50

30-Inch Focal Distance Aligned Grids

10" x 12"	\$ 95.00
11" x 14"	110.00
14" x 17"	137.50

DEVELOPING TANKS

14" x 36" Powell Non-refrigerated Five Gallon Capacity Without Cover.....	\$133.00
Stainless Steel Cover for Above Tank	12.50
14" x 17" Powell Non-refrigerated Five Gallon Capacity Without Cover.....	79.50
Stainless Steel Cover for Above Tank	8.25
14" x 17" Powell Refrigerated Five Gallon Capacity Including Stainless Steel Cover.....	435.00
14" x 17" Powell Five Gallon Capacity Improved Non-refrigerated Tank Without Cover.....	110.00
Stainless Steel Cover for Above Tank	11.75
Powell Dryer for Films Up to and Including 14" x 17" With Cover.....	295.00
14" x 17" Powell Water Cooled Five Gallon Capacity Including Cover.....	467.50
14" x 17" Wolf Five Gallon Capacity Tank.....	60.00
8" x 10" Wolf One Gallon Capacity Tank.....	35.00
14" x 17" Wolf Two-and-One-Half Gallon Tank.....	50.00

DEVELOPING TANKS (CONT'D)

5" x 7" Wolf One-Half Gallon Capacity tank.....	\$ 15.00
Kaye Self-Contained Refrigerated Five Gallon Capacity Tank.....	485.00
Buck Five Gallon Capacity Tank.....	55.00
Buck Two-and-One-Half Gallon Capacity Tank.....	44.00
Buck Dental Tank One Gallon Capacity.....	28.50
Buck Dental Tank With Overflow Stand Pipe.....	36.20

"HAIRPIN" PORTABLE IMMERSION HEATER.....	\$ 13.50
--	----------

THERMOMETERS

Combination Thermometer and Stirring Device.....	\$ 1.85
Floating Thermometer.....	2.00

FILM STORAGE CABINETS

All Metal Film Chest 8 1/8" x 19 5/8" x 19" Hammered Grey Finish.....	\$ 32.00
All Steel 14" x 36" Cabinet With Cover Hammered Grey Finish.....	80.00

ILLUMINATORS

Fischer Powell 14" x 36" Illuminator.....	\$ 55.00
Buck 14" x 17" Illuminator.....	19.80
Buck Fluorotube 14" x 17" Illuminator.....	32.50
Fischer 14" x 36" Illuminator.....	75.00
Paragon Economy Illuminator 14" x 17".....	15.00
Paragon Dental Illuminator.....	8.00

INTENSIFYING SCREENS

Buck X-TRA Speed and Buck MID Speed Screens

5" x 7"\$ 3.20 - Pair	11" x 14"\$ 21.00 -Pair
6 1/2" x 8 1/2" 6.50 - Pair	14" x 17" 32.00 -Pair
8" x 10" 10.50 - Pair	14" x 36" 67.50 -Pair
10" x 12" 16.00 - Pair	

Eastman Ultra Speed Screens

5" x 7"\$ 6.50 - Pair	11" x 14"\$ 28.40 -Pair
6 1/2" x 8 1/2" 11.60 - Pair	14" x 17" 44.00 -Pair
8" x 10" 14.80 - Pair	14" x 36" 93.10 -Pair
10" x 12" 22.20 - Pair	

Patterson High Speed Screens

5" x 7"\$ 6.60 - Pair	11" x 14"\$ 29.30 -Pair
6 1/2" x 8 1/2" 12.00 - Pair	14" x 17" 45.20 -Pair
8" x 10" 15.20 - Pair	14" x 36" 95.80 -Pair
10" x 12" 22.80 - Pair	

Patterson Par Speed Screens

5" x 7"\$ 5.90 - Pair	11" x 14"\$ 25.85 -Pair
6 1/2" x 8 1/2" 10.50 - Pair	14" x 17" 39.90 -Pair
8" x 10" 13.45 - Pair	14" x 36" 84.65 -Pair
10" x 12" 20.15 - Pair	

7" x 7"	\$ 11.00
8" x 8"	13.20
9" x 10"	14.85
10" x 12"	17.05

11" x 14"	\$ 19.25
14" x 17"	21.45
14" x 36"	49.50

X-RAY FILMS

25 SHEET PKG.

2" x 7"	3.45
2 1/2" x 8 1/2"	5.25
3" x 10"	7.35
4" x 12"	10.80
5" x 14"	13.80
6" x 17"	21.10

75 SHEET PKG.

.....	\$ 10.05
.....	15.30
.....	21.40
.....	31.45
.....	40.15
.....	61.40

25 SHEET PKG.

5" x 7"	\$ 2.85
6 $\frac{1}{2}$ " x 8 $\frac{1}{2}$ "	4.35
8" x 10"	6.10
10" x 12"	8.95
11" x 14"	11.45
14" x 17"	17.50
14" x 36"	36.60

75 SHEET PKG.

.....	\$ 8.35
.....	12.70
.....	17.75
.....	26.10
.....	33.30
.....	50.95
.....	Not Stocked

14" x 36".....	38.80	
<u>KODAK X-RAY FILM CORNER CUTTER.....</u>		\$ 11.00

X-RAY EXPOSURE HOLDERS

Kodak	5" x 7"	\$.40
Kodak	6 $\frac{1}{2}$ " x 8 $\frac{1}{2}$ "55
Kodak	8" x 10"55

Kodak	10" x 12"	\$.65
Kodak	11" x 14"80
Kodak	14" x 17"95

NEGATIVE PRESERVERS (FILING ENVELOPES)

Manila Filing Envelopes

5" x 7".....	\$ 1.40 per 100
6½" x 8½".....	1.80 per 100
8" x 10".....	1.80 per 100

10" x 12".....	2.65 per 100
11" x 14".....	3.00 per 100
14" x 17".....	3.80 per 100

DEVELOPING HANGERS

Fischer 14"x36" Fold Over Developing Hanger for Use in 14"x17" 5 Gallon Tank	\$ 7.50
Powell 14" x 36" Hanger	8.00
Fischer 5" x 7" Hanger	2.00
Fischer 6 1/2" x 8 1/2" Hanger.....	2.10
Fischer 8" x 10" Hanger	2.10
Fischer 10" x 12" Hanger.....	2.15
Fischer 11" x 14" Hanger.....	2.20
Fischer 14" x 17" Hanger.....	2.25

FILM DRYING RACK

Paragon Drying Rack	\$ 2.50
---------------------------	---------

DEVELOPER

	Powder Form	
Kodak and Buck X-Ray Developer Sufficient to Prepare 1 Gallon Solution.....	\$.85
Kodak and Buck X-Ray Developer Sufficient to Prepare 5 Gallon Solution.....		3.50

Concentrated (Liquid) Form

Kodak and Buck X-Ray Developer Sufficient to Prepare 1 Gallon Solution.....	\$	1.25
Ansco Developer Sufficient to Prepare 1 Gallon Solution		1.25
Kodak, Buck and Ansco X-Ray Developer Sufficient to Prepare 5 Gallon Solution		4.50

FIXER

	Powder Form	
Buck Fixer Sufficient to Prepare 1 Gallon Solution.....	\$.85
Kodak Fixer Sufficient to Prepare 1 Gallon Solution.....		.85
Buck Fixer Sufficient to Prepare 5 Gallon Solution.....		3.00
Kodak Fixer Sufficient to Prepare 5 Gallon Solution.....		3.25

Concentrated (Liquid) Form

Kodak, Buck and Ansco Fixers Sufficient to Prepare 1 Gallon Solution.....	\$	1.25
Kodak, Buck and Ansco Fixers Sufficient to Prepare 5 Gallon Solution.....		4.50

THICKNESS-OF-PART CALIPER

Fischer Thickness-Of-Part Caliper.....	\$	2.50
--	----	------

X-RAY FILM FILING CABINETS

Fischer - Bentson

	Olive Green or Grey	Walnut or Mahogany
Three Large Drawers with Small Temporary Flat Filing Drawer At Top	\$ 96.50	\$ 113.00
Same with General Lock	108.00	124.50
Three Drawer Cabinet Without Temporary Drawer.....	89.00	103.25
Same with General Lock.....	100.50	114.75
Two Drawer Cabinet With Small Temporary Flat Filing Drawer at Top.....	80.50	94.25
Same with General Lock	92.00	105.75
Two Drawer Cabinet Without Temporary Drawer.....	73.25	84.75
Same with General Lock	84.75	96.25

X-RAY FILM FILING CABINETS (CONT'D)

- 5 -

Fischer - Bentson

Single Drawer Cabinet.....	Olive Green	Walnut or
Same with Individual Lock.....	or Grey	Mahogany
	\$ 40.00	\$ 48.50
	44.25	52.75

DARK ROOM LAMPS

Kodak Lamp with 5½" Safelight.....	\$ 4.67
Kodak Adjustable Lamp with 5½" Safelight.....	7.58
Brownie Lamp with Rectangular and Circular Safelight.....	3.79
Buck Lamp 8" x 10" Fitted with Safelight.....	16.80

GASTROPAQUE

20 Dose Package (Flavored).....	\$ 2.25
10 Pound Package (Flavored).....	2.50
10 Pound Package (Unflavored).....	2.00

GALL BLADDER OPAQUE

Paragon Gall Bladder Opaque Single Bottle.....	\$.90
Paragon Gall Bladder Opaque One Dozen Bottles.....	4.60

BARIUM MEAL

5 Pound Can.....	\$ 1.50
10 Pound Can.....	2.50
25 Pound Can.....	5.75
100 Pound Can.....	20.00

NEO-IOPAX

20 cc. Ampule.....	\$ 2.00
--------------------	---------

TETRADOL EMULSION

Single Dose.....	\$ 1.25
One Dozen Doses.....	9.70

DIODRAST

20 cc. Ampules, Ready for Use.....	\$ 1.75
20 cc. Ampules, Box of Ten.....	15.35

LEADED RUBBER GLOVETS

Lead Rubber Glovets Pair.....	\$ 14.00
Left Hand Glovet Only.....	8.40

LEATHER-COVERED LEADED RUBBER X-RAY PROTECTION GLOVES

Medium Size.....	\$ 15.50
------------------	----------

X-RAY APRONS

Waterproof Apron.....	\$ 2.50
Leaded Rubber Apron.....	15.00

TIMERS

Hawkeye Measured Timer - Spring Wound.....	\$ 9.50
Model "S" Minute Minder - 60 Minute Range.....	6.50

X-RAY FILM MARKERS

Perfection Marker with Four Sets of 1/4" Figures and One Each R and L In Wood Case....	\$ 7.00
Set of Months and Dates.....	3.75
Perfection Marker Only.....	2.75
Eclipse Cushion Film Marker Including:	
1 Marker with Name and Address.....)	
4 Each 1/4" Figures 0 to 9 Mounted)	
1 Each 1/4" Letters R and L Mounted)	\$ 7.50
1 Eighteen Compartment Case.....)	
1 Months and Dates Set	\$ 2.50
Figures Mounted..... Each	.06
Marker Only Stenciled.....	3.00
Ideal Cushion Marker Including:	
1 Marker Stenciled.....)	
4 Each 3/16" Figures 0 to 9 Mounted)	
1 Each 3/16" Letters R and L Mounted)	\$ 11.50
2 Each 3/16" Letters A to Z Mounted)	
1 Each 3/16" Letters A-E-I-O-U.....)	
1 36 Compartment Case.....)	
1 Months and Dates Set.....	\$ 3.75
1 Months and Dates Set In Figures Only.....	2.50
1 Marker Only Stenciled.....	3.25
Figures or letters Mounted.....	.06
Wheel Marker with Month and Date Attachment:	
Marker with Month and Date Attachment.....	\$ 10.00
Attachment only.....	3.00
Marker Only.....	7.00
Lead Letters and Figures for Above Markers:	
Mounted.....	.06
Unmounted.....Each	.04

X-RAY GLASSES

Paragon Fluoroscopic Goggles With Green Lense Only.....	\$ 2.75
Paragon Protective Fluoroscopic Goggles With Lead Glass and Dark Olive Green Lenses.....	6.50
Paragon Fluoroscopic Glasses.....	3.50
Paragon (Non-Protective) Fluoroscopic Goggles - Red or Green.....	2.50

United States Navy



BUREAU OF MEDICINE AND SURGERY

Certificate of Achievement

to

H. G. Fischer & Co.

In grateful acknowledgment of exceptional cooperation and outstanding services rendered to the Medical Department of the Navy during the period of World War II

A. H. Dearing

Rear Admiral, MC, U. S. Navy
Surgeon General of the Navy



H. G. FISCHER & CO. FRANKLIN PARK, ILLINOIS

Manufacturers of high quality X-ray and electromedical equipment since 1910.

Holders of the U. S. Army-Navy "E" Award with three additional awards represented by the stars on the flag, and holders of the U. S. Navy Certificate of Achievement — all for outstanding services rendered.


This series of awards constitutes the most signal recognition granted any manufacturer of x-ray equipment.

Producing — in a new 82,000 sq. ft. factory — efficient, durable, attractive apparatus, which offers customers the greatest value per dollar expended obtainable in the entire industry.



A. H. Dearing, Rear Admiral, MC, U. S. Navy, presenting Certificate of Achievement to A. W. Mathis, President, H. G. Fischer & Co. (Official U. S. Navy photograph.)

OVER 100,000 SATISFIED USERS THROUGHOUT THE WORLD



NEW HOME OF H. G. FISCHER & CO.

Representative Units of the Complete Line of H. G. FISCHER & CO.

The complete H. G. FISCHER & CO. line includes many models of Shockproof X-ray apparatus (medical and dental), Short Wave Diathermy Units, Galvanic and Contractile Currents Generators, Ultra Violet Generators, Infra Red Generators, other apparatus, supplies and accessories. H. G. FISCHER & CO. has always maintained the policy of meeting or exceeding the standards of any competing units designed for the same purpose, regardless of price, and pricing all apparatus at the lowest point consistent with quality manufacture.

SINCE 1910, H. G. FISCHER & CO. has been one of the largest manufacturers of Shockproof X-ray and Electromedical equipment. Pioneers in the field of X-ray and Electromedical equipment, H. G. FISCHER & CO. is recognized throughout the world today for apparatus of dependable service and great durability. H. G. FISCHER & CO. is the ORIGINAL COMPANY. It has no connection with any other company having a similar name. No H. G. FISCHER & CO. equipment is manufactured outside of the Chicago area.

More than thirty-eight years of experience in engineering, design and construction of X-ray and Electromedical apparatus, together with fine workmanship and materials, are built into all H. G. FISCHER & CO. apparatus. The fact that today over 100,000 physicians, dentists, hospitals, clinics, universities and other medical organizations in the United States and throughout the world use H. G. FISCHER & CO. apparatus reflects its outstanding quality and service.

All H. G. FISCHER & CO. apparatus is serviced and sold through factory-trained representatives who are well qualified to render necessary technical assistance. Our comparatively small overhead expense and factory-trained representation enable us to give purchasers such value per dollar expended as cannot be excelled by any other manufacturer in the field.

The new home of H. G. FISCHER & CO. is now conveniently located at 9451-91 W. Belmont Ave., Franklin Park, Illinois, a western suburb of Chicago (3200 North and 9400 West, by the Chicago numbering system). Doctors are cordially invited to visit the new H. G. FISCHER & CO. plant, where modern X-ray and Electromedical equipment will be displayed and demonstrated. A visit entails no obligation.

H. G. FISCHER & CO.

9451-91 W. BELMONT AVE. • FRANKLIN PARK, ILLINOIS



De Luxe RF 250 Milliamperes X-ray Apparatus with Hand-Tilted or Motor-Driven Table.



De Luxe RF 100 Milliamperes X-ray Apparatus with Hand-Tilted or Motor-Driven Table.



Model MPA 20 or 30 Milliamperes Mobile-Portable Shockproof X-ray Apparatus.



100 Milliamperes X-ray Unit with Urological Table.



Table Model "200" Short Wave Diathermy Unit.



De Luxe Model "400" Short Wave Diathermy Unit.



Spacesaver Radiographic-Fluoroscope Unit and Examining Table, 30, 50, 100, 250 Milliamperes.



Model TC 30 or 50 Milliamperes Shockproof X-ray Apparatus.



Model TF 30 or 50 Milliamperes Shockproof Fluoroscope with Radiographic Facilities.



Mobile Shockproof Dental X-ray.



Model "J" Galvanic and Contractile Currents Generator.



Model "CK" Ultra Violet Generator.

Printed in U.S.A.

350 Form 3125A